COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 326**
IEC 61010-1:1990+A1:1992+A2:1995	14.7	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Mains transformers	- Temperature - Melting - Limits	

Question:

What temperature limits should be applied in single fault condition?

Decision:

Refer to Table4 of IEC61010-1 'Motor temperature'.
Note: Where melting of insulation occurs this is not significant where it does not cause a hazard.

Explanation:

Table 4 is to be amended to include transformers in its application in the next edition and a text change included referring to hazards when insulation melts.