COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 327**
IEC 61010-1:1990+A1:1992+A2:1995	12	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Protection against Light source Radiation	- Light Source - Radiation	

Question:

What limits have to be applied to radiation from bright light and Ultra-Violet Radiation – sources?

Decision:

Currently limits are specified not in IEC61010-1, however local regulations apply in some countries. One could refer to IEC 60825 and IEC 60598-1:1996 clause 4.24.

Explanation: ---