COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 331**
IEC 61010-1:1990 +A1:1992 + A2 :1995	1.4	**Page 1(1)**
Subject	**Key words:**	**Decision n. 23 taken at 37th meeting/2000**
Environmental conditions	- Higher humidity - Higher temperature	

Question:

When equipment is rated for higher humidity / temperature levels than those in this clause, how is this considered for evaluation?

Decision:

See and follow Explanatory Notes below.

Explanatory Notes:

Note 1 under sub-clause 4.1 suggest that additional tests should be considered to cover the full range of the environment conditions where this may have a bearing on safety.

Note: See also the new sub-clause 1.4.2 Extended environmental conditions, 2nd Edition IEC61010-1.