COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 332**
IEC 61010-1:1990 +A1:1992 + A2 :1995	6.5	**Page 1(1)**
Subject:	**Key words:**	**Decision n. 23 taken at 37th meeting/2000**
Protection in single fault condition	- Motor controllers - Protective earthing	

Question:

Certain motor controllers have difficulty meeting 0.5mA under normal condition and 3.5mA in SFC. Is it possible to have two ways of permanent protective grounding so that in SFC still connection to PE is assured (for permanent connections only).

Decision:

Follow IEC61010-1 requirements.

Explanatory Notes:

The requirement for protective earthing is described under sub-clause 6.5. It should be noted that under normal and single fault conditions accessible parts shall meet requirements.

For permanently connected equipment the protective earth must not be disconnected as a single fault.