COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 333**
IEC 61010-1:1990 +A1:1992 + A2 :1995	6.6.1	**Page 1(1)**
Subject:	**Key words:**	**Decision n. 23 taken at 37th meeting/2000**
Separation of Internal circuits	- Isolation system - Materials Declaration	

Question:

What should be the criteria for accepting a manufacturer's declaration for isolation systems and materials?

This question is from CTL.

Decision:

It was agreed that although manufacturer 's declaration may well be taken into consideration, however testing must be performed to ensure that requirements of IEC61010-1 is met.

Explanatory Notes:

In IEC61010-1, there is no requirement to simulate the ageing process of critical components such as the mains / isolating transformers.

It should be noted that after each single fault condition test, where necessary, the required high voltage test must also be applied to ensure that at least basic insulation exists.

Note that under compliance criteria there is no reference to 'Declaration by manufacturer'.