COLLECTION OF CTL DECISIONS

Standard: IEC 61010-1:1990 +A1:1992 + A2 :1995	**Sub clause:** 10.1	**Sheet n. 337** **Page 1(1)**
Subject: Integrity of clearances and creepage distances	**Key words:** - Clearances distances - Creepage distances	**Decision n. 23 taken at 37th meeting/2000**

Question:

Strict interpretation does implicate a lot of test work.
How is this clause considered for compliance?
This question is from CTL.

Decision:

See Explanatory Notes

Explanatory Notes:

The compliance requirements suggest that, in case of doubt additional tests are required to ensure the integrity. Where examination of the data and the test results cause concern then further tests are needed at 40°C ambient followed by measurements of the concerned area for the creepage and clearance distances.