COLLECTION OF CTL DECISIONS

Standard: IEC 61010-1:1990 +A1:1992 + A2 :1995	**Sub clause:** 14.3	**Sheet n. 338** **Page 1(1)**
Subject: Overtemperature protection devices	**Key words:** Re-settable Cut-outs Re-settable Circuit Breakers	**Decision n. 23 taken at 37th meeting/2000**

Question:

This clause describes requirements for overtemperature protective devices operated in a single fault condition. What about the requirements for overcurrent protective device, such as manual re-settable type circuit breakers?

Decision:

Appropriate approval data must be available for examination and where necessary additional tests are to be conducted to ensure compliance with IEC 61010-1.

Explanatory Notes:

According to sub-clause 14.1, where safety is involved, critical components must comply with the relevant IEC standards. It is required that such components be suitably rated and appropriate for the intended function.