COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 343**
IEC 61010-1:1990 +A1:1992 + A2 :1995	4.3.14	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Loading and filling	- Loading and filling - Water substitute	

Question:

Can water be substituted for loading and filling purposes where appropriate?

Decision:

It was agreed that water may be substituted for most fluids. Where leakage or spillage occurs then, if necessary to determine the consequences, the actual fluid may be applied in limited locations.

In case of doubt, the specified material should be used.

Explanatory Notes:

-