COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 345**
IEC 61010-1:1990 +A1:1992 + A2 :1995	5.1.4	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Fuses (External markings)	Markings for external fuses	

Question:

What do Test Houses generally require for compliance to this sub-clause?

Decision:

Sufficient information should be given such that the operator/user can replace the fuse with the correct one. This information should include the rated current, rupture capacity (high or low), the speed and the voltage rating either in symbols or words.

There is no requirement for the use of HRC fuses in the supply lines.(see also clause 9.6) The markings and documents should make it clear whether an IEC60127 or UL/CSA type fuse is fitted.

Explanatory Notes:

BSI/VDE/BREZ/KEMA require an IEC 60127 style fuses if operator / user replaceable, however, UL only accepts UL/CSA listed / certified fuses for USA sales, this includes UL recognised IEC 60127 fuses.