COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 346**
IEC 61010-1:1990 +A1:1992 + A2 :1995	6.10	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Connection to mains supply source	- Mains supply cords - Cross - section area Dimensions	

Question:

In the absence of a Table for assessing compliance for the size of the mains supply cords, what should be the criteria for compliance?

Decision:

It was agreed that Table VIII 'sizes of conductors in power supply cords' of IEC 60950 should be referenced.

Explanatory Notes:

The Table referred is used in other IEC Standards comparable to IEC 61010-1.