# COLLECTION OF CTL DECISIONS

| **Standard:** | **Sub clause:** | **Sheet n. 347** |
|---|---|---|
| IEC 61010-1:1990 +A1:1992 + A2 :1995 | 6.10.2 | **Page 1(1)** |
| **Subject:** | **Key words:** | **Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto** |
| Fitting of non-detachable mains supply cords | - Fixed equipment<br>- Cable clamp | |

**Question:**
For fixed equipment and where the mains cord is accessible (not part of the building wiring), is the use of a cable clamp necessary for the construction?

**Decision:**

It was agreed that if the non-detachable mains supply cord is accessible, even if the equipment is fixed, then a cable clamp should be provided since the cable itself could be subject to movement.

**Explanatory Notes**:

-