COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 348**
IEC 61010-1:1990 +A1:1992 + A2 :1995	6.10.3	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Plugs and connectors	Plug pins charge shock hazard	

Question:

How is the measurement under 6.10.3 c) performed?

Decision:

It was agreed that the test should be repeated to capture the peak value.

Explanatory Notes:

-