COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 349**
IEC 61010-1:1990 +A1:1992 + A2 :1995	6.12.3.1 Switches and circuit breakers	**Page 1(1)**
Subject;	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Switches and circuit breakers	- Disconnecting devices - Switches - Circuit breakers	

Question:

Which are considered to be the relevant requirements of IEC 60947-1?

Decision:

It was agreed that the relevant requirements include adequacy of rating for the load and life testing e.g. 10 000 cycles of operation.

Explanatory Notes:

-