COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 350**
IEC 61010-1:1990 +A1:1992 + A2 :1995	8	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Mechanical resistance to shock, vibration and impact	Immersion heaters Enclosures	

Question:

Should the mechanical resistance to shock and impact test be performed on devices manufactured with glass?

Decision:

It was agreed that this clause applies to the glass enclosures of immersion heaters, for example.

Explanatory Notes:

-