COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 351**
IEC 61010-1:1990 +A1:1992 + A2 :1995	9.3	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Guards	- Guards - Hot surfaces	

Question:

Which surfaces are allowed to be hot under this sub-clause?

Decision:

It was agreed that surfaces, which exceed the temperatures, specified in Table 3 are permitted if they are marked as required by this clause.

Explanatory Notes:

-