COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 352**
IEC 61010-1:1990 +A1:1992 + A2 :1995	9.6	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Overcurrent protection	- Fuse types - High Rupture Capacity (HRC)	

Question:

Do laboratories place any limitation on the type and size of fuse which may be used in the mains supply connections?

Decision:

It was agreed that there are no requirements that fuses used in the supply lines be of high rupture capacity. They should, however, be suitable for their intended function.

Explanatory Notes:

-