COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 354**
IEC 61010-1:1990 +A1:1992 + A2 :1995	12.5.2	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Ultrasonic pressure	Ultrasonic pressure Limits	

Question:

What should be the limit for this requirement?

Decision:

It was agreed that engineering judgement should be used to determine if the test is necessary and that the National limits apply.

Explanatory Notes:

-