COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 355**
IEC 61010-1:1990 +A1:1992 + A2 :1995	13.1	**Page 1(1)**
Subject:	**Key words:**	**Provisional decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Poisonous and injurious gases	- Poisonous gases - Injurious gases - Dangerous amounts	

Question:

1. What criteria do laboratories employ to determine acceptable limits for gases?

2. What techniques do laboratories employ to detect and identify dangerous gases?

Decision:

It was agreed that National Occupational Threshold Limits apply. The limits apply to gases released both during normal and single fault conditions.

NOTE: It was mentioned that this sub-clause may not appear in the Second Edition of IEC 61010-1.

Explanatory Notes:

-