COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 356**
IEC 61010-1:1990 +A1:1992 + A2 :1995	13.2.2	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Batteries	- Explosion - Fire hazard - Charge state	

Question:

The state of charge of the battery will affect the result obtained. How do laboratories determine the initial charge condition?

Decision:

It was agreed that the battery data should be assessed to establish the worst case condition and that the battery should be arranged to be in that condition.

Explanatory Notes:

-