COLLECTION OF CTL DECISIONS

Standard: IEC 61010-1:1990 +A1:1992 + A2 :1995	**Sub clause:** 14.2.1	**Sheet n. 357** **Page 1(1)**
Subject: Motor temperatures	**Key words:** - Overtemperature protection devices - Self-resetting devices	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**

Question:

How should motor temperatures be monitored when fitted with a self-resetting temperature device?

Decision:

It was agreed that temperatures should be monitored during the whole 200 cycle (see SC14.3) period and that thermocouples may be used to monitor the winding temperature.

The thermocouples should be located such that they do not interfere with the test.

Explanatory Notes:

-