COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 358**
IEC 61010-1:1990 +A1:1992 + A2 :1995	14.6	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
High integrity components	- High integrity - Reliability - Fault free	

Question:

Is there any listing of these devices and rules for their use?

Decision:

The laboratories were not aware of any comprehensive listing of these devices. It was agreed that examples of these devices include EMI Y capacitors (IEC 60384-14) and fuses (IEC 60127).

NB; PTCs are not considered to be a high integrity component.

Explanatory Notes:

-