COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 360**
IEC 61010-1:1990 +A1:1992 + A2 :1995	Annex D	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Application of Annex D	Creepage & clearance distances	

Question:

How do Test Houses apply Annex D.5.2?

Decision:

It was agreed that the result of calculations to Section D.5.2 may result in clearance distance requirements that are less than those specified in Table D1 to D10.

Explanatory Notes:

-