COLLECTION OF CTL DECISIONS

Standard: IEC 61010-1:1990 +A1:1992 + A2 :1995	**Sub clause:** 4.4	**Sheet n. 361** **Page 1(1)**
Subject: Testing in single fault condition	**Key words:** - Voltage selector - Voltage selecting mismatch	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**

Question:

For products employing a voltage selector switch (120/230V) for an input voltage which is accessible to the user, should a mismatch abnormal test be considered?

Decision:

SFC test must be considered as required by sub-clause 4.4.1 where there is a hazard.

Explanatory Notes:

Note that for clarity, CSA is compiling text which will than be put to WG 1 TC66 as proposal for possible inclusion in the 1st Amendment to IEC61010-1 2nd edition.