COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 363**
IEC 61010-1:1990 +A1:1992 + A2 :1995	6.5.1.3	**Page 1(1)**
Subject:	**Key words:**	**decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Bonding impedance of P.C.E	- Mains cord - Impedance	

Question:

1. This clause specifies that the mains cord impedance does not form part of the specified bonding impedance. Dose this apply to cord connected products? If so, what about the impedance of the cord?

2. Should the resistance of interconnection cable between units of a system be disregarded like mains cable?

Decision:

1.It was decided that the supply mains cord must be included in the measurement and that the total limit must not exceed 0.2Ω.

2. Furthermore, other interconnection cables (if any) must also be included in the measurements.

Explanatory Notes:
-