COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 366**
IEC 61010-1:1990 +A1:1992 + A2 :1995	6.10.1	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Mains supply cords	- Mains cord - Temperature	

Question:

1. No requirements are specified for the suitability of mains cords when the range of Certified temperatures (max 250°C as far as I know) are exceeded 550°C surface temperature for example have been measured in a test.

2. The temperature rating of the appliance coupler is not based on temperature of the heating equipment that is able to generate, it is based on temperature of the contacts. Therefore there is no relation in the way as it is written in IEC61010-1 standard.

Decision:

1. Adequate provision must be provided in the construction of the product and in the associated user handbook to ensure that the cord used does not exceed the maximum temperature it is certified for under normal/sfc. For example, protection of the cord is required in particular at the point of entry to the equipment, as well as the parts where it is liable to be damaged in use.

2. In such extreme cases it is suggested that the equipment be fitted with a non-detachable cable in line with the decision 1 above and not a detachable supply cord.

Explanatory Notes:

-