COLLECTION OF CTL DECISIONS

Standard: IEC 61010-1:1990 +A1:1992 + A2 :1995	**Sub clause:** 9.2	**Sheet n. 367** **Page 1(1)**
Subject: Temperature tests	**Key words:** - Internal wiring - Size/Capacity - High temperature	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**

Question:

For internal wiring, in particular wiring near to heat source or getting hot - what should be the criteria for compliance (size & temperature requirements)?

Decision:

If the wiring is regarded as safety critical component than they must be rated adequately and assessed accordingly, i.e. following tests, inspect the wiring for any damage and if in doubt apply the required HV test and check temperature test results with the rating/information supplied.

Explanatory Notes:

-