COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 368**
IEC 61010-1:1990 +A1:1992 + A2 :1995	14.7.2/4.4.2.6	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Overload test	- Mains transformer - SFC tests	

Question:

What criteria should be used when deciding whether or not to do the overload test as there seems to be subjectivity involved when considering the requirements?

Decision:

It was decided that transformers should be considered for SFC tests (short circuit and overload tests) under the worst possible scenario.

Explanatory Notes:

-