COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 369**
IEC 61010-1:1990 +A1:1992 + A2 :1995	Tables in Annex D	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Comparative Tracking Index	- CTI values - Manufacturer's claim	

Questions:

1. Is the manufacturer's declaration regarding the CTI values acceptable by Test Houses?

2.Is a single test for the CTI value acceptable?

3.Which CTI value is to be used when no value tested or declared, for example for PCB's?

Decision:

1. The manufacturer's claim is not adequate and must be backed by an independent test report/certificate from a recognised test house.
2. No testing considered necessary if the material is approved and fit for purpose. However if not approved than single test is not acceptable.
3. The least favourable value, for example CTI value of < 100.

Explanatory Notes:

-