COLLECTION OF THE CTL DECISIONS
DECISION SHEET

Standard(s)- (year and edition): IEC 61010-1:2002 Ed.2	**Sub clause(s):** 6.7.2	**Sheet n°: 536**
Subject: CAT III for mains circuits	**Key words:** Mains, CAT III	Decision taken at the 40th meeting 2003

Question:

Measuring equipment (not kWh-meter) that is permanently installed in the service equipment has a power supply unit which is connected close to the incoming service and the kWh-meter. The device sends information to the electrical company, utilizing the mains service and connected close to it. This is to my knowledge a CAT III area.

Since CAT III has been withdrawn for mains circuits (6.7.2), the requirement for a 300 V mains would be 3 mm air / 3-6 mm creepage / 2 224 Vrms Test voltage on the supply side, according to table 4 and 9.

Decision:

In this case you can apply CAT III for measuring circuits.

Explanatory notes:

This issue has been acknowledged by TC66 and corrected in future standard publications.