CTL DECISION SHEET

Standard(s)- (year and edition):	Sub clause(s):	Sheet n°: DSH 618
IEC 61010-1, ed. 2, 2001	9.2.1 a)	
Subject:	**Key words:**	**Decision approved by the CTL Plenary meeting 2007**
Flammability rating of internal wiring	Wires, flammability,	

Question:
Insulated wires shall be equivalent FV-1 or better. Wires are not generally tested and rated according to FV-1. It is not easy to establish equivalence between different test classifications/methods if you are not involved directly in such testing procedures.
Wires typically are rated VW-1 or FT-1. It is unclear if these flammability ratings are equivalent and thus acceptable (Same problem exists with IEC 60601-1 ed. 3)

Decision:

Wires rated e.g. VW-1 or FT-1 or according to IEC 60332-1-2 are deemed equivalent and can be accepted.

Explanatory notes:

IECEE Secretariat, 3, rue de Varambè Geneva, Switzerland
Telefax: + 41 22 919 0300, e-mail: pro@iec.ch