CTL DECISION SHEET

Standard(s)- (year and edition):	Sub clause(s):	Sheet n°: DSH 619
IEC 61010-1, ed. 2, 2001	-	
Subject:	**Key words:**	**Decision approved by the CTL Plenary meeting 2007**
3 Phase power supply	Requirement; power supply	

Question:

Dec. 35: 43/06 of the 43th CTL meeting: The CTL agreed that all ETFs should examine the power supply requirements in standards (e.g. need for a 3 -phase power supply) and update the lists of test equipment to include these requirements, where applicable.

Decision:
List has been updated; for the main part a 3 phase power supply does not need to be present at the CBTL, but can be used in MTL situations (status "W")

Explanatory notes:
The status "S" is not applicable. If the instrument under test requires a 3 phase power supply, a majority of tests need to be performed (subcontracted) outside the CBTL, which is not allowed.

IECEE Secretariat, 3, rue de Varambè Geneva, Switzerland
Telefax: + 41 22 919 0300, e-mail: pro@iec.ch