CTL PROVISIONAL DECISION SHEET (PDSH)

Standard(s) (incl. year)	Subclause(s)	No.	Year
IEC 61010-1:2011	14.1	PDSH 2016	2015
Category			
MEAS			
Subject	**Keywords**	**Developed by**	**To be approved**
Acceptance of Transformers certified to IEC 61558 for use in IEC61010-1 equipment	Transformers, IEC 61558, Component acceptance	ETF-3	2016 CTL Plenary Meeting

Question

IEC 61010-1, clause 6.7.2.1 (Table 4) requires a minimum creepage distance between primary and secondary circuits of 3.0 mm (pollution degree 1, up to 300 Vrms)

A potted transformer certified to IEC 61558-1, IEC 61558-2-1provides only 1.4 mm at the same area, which is within the limits of the IEC 61558 standards.

Is it acceptable to use such a transformer as measuring equipment (portable appliance tester) under consideration of the requirements of clause 14.1?

Decision

No, the component has to fulfil the insulation requirements of IEC 61010-1.

Explanatory notes

© IEC – IECEE 2013