☑ CTL Provisional DECISION SHEET (PDSH)
☐ CTL DECISION SHEET (DSH)

Standard(s) (incl. year)	**Subclause(s)**	**Tracking No.**
IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016	1.4 4.3.2.1 10.	2189A
Category		
MEAS		
Subject	**Keywords**	**Developed by**
Acceptable derating regarding Cl. 1.4 and 4.3.2.1	Derating, Input, Temperature Test	ETF 3

Question

Is it possible to have an output derating within the voltage fluctuation of the mains supply?

Example:

Input: 100-240Vac (+10%, -15%)

(Input <100Vac derating the output 1%/V)

Decision

No, it is not acceptable to derate the output A/W/VA, as the input voltage drops within the specific tolerance of the manufacturer.

The standard requires compliance within all tests at the tolerance stated in 1.4.1 or the wider tolerance stated by the manufacturer.

The temperature test shall comply with the input voltage range noted in 1.4.1.

Explanatory notes

Enquiry:	2189
Date:	2022
PDSH:	2189A
Year of publication:	2025
To be approved at CTL plenary:	2026

DSH

Publication date:

Click here to enter text.

Disclaimer: This document is controlled and has been released electronically.
Only the version on the IECEE Website is the current document version.