☑ CTL Provisional DECISION SHEET (PDSH)
☐ CTL DECISION SHEET (DSH)

Standard(s) (incl. year)	**Subclause(s)**	**Tracking No.**
IEC 61010-1:2010 IEC 61010-1:2010/AMD1:2016	4.4.2.5	2291A
Category		
MEAS		
Subject	**Keywords**	**Developed by**
Temperatures of windings table 20 SFC	SFC, Temperatures, Motors	ETF 3

Question

In contrast to the limits in IEC 60601-1, IEC 60335 and IEC 60085, Table 20 in IEC 61010-1:2010+A1:2016, which references IEC 60085, indicates only one temperature limit for each class.

Is this limit the arithmetic average after the first hour?

Decision

No, the limits in Table 20 shall be considered the maximum insulation temperature during testing.

It is recognized that other standards define the measured limits over time in single fault conditions different from IEC 61010-1:2010 + A1:2016. Any changes to the requirements within the standard must be addressed by TC66 WG1. A developed proposal should be submitted to TC66 WG1 for consideration in a future revision.

Explanatory notes

Enquiry:	2291
Date:	2025
PDSH:	2291A
Year of publication:	2025
To be approved at CTL plenary:	2026

DSH

Publication date:

N/A

Disclaimer: This document is controlled and has been released electronically.
Only the version on the IECEE Website is the current document version.