COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 319**
IEC 61010-1990+A1:1992+A2:1995	4.4.4.1	**Page 1(1)**
Subject:	**Key words:**	**Decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Compliance for requirements for electric shock protection	- Shock - Fault condition	

Question:

If a fuse operates when checking equipment in Single Fault Condition as required by Clause 6.3.2, should the fuse be replaced before the measurement of voltage, current and capacitance are made?

Decision:

Where the fuse link is operator replaceable it should be replaced up to 3 times to enable the measurements of 6.3.2 to be performed.

Explanation:

Where operators are instructed to replace a fuse link this should be done for a maximum of 3 times and no hazard as defined in 6.3.2 should arise.

Where the link is not designated as replaceable by the Operator then this should not be done.