COLLECTION OF CTL DECISIONS

Standard:	**Sub clause:**	**Sheet n. 309**
IEC 61010	9.6	**Page 1(1)**
Subject:	**Key words:**	**Provisional decision taken by ETF3 and confirmed by CTL at its 38th meeting, in Toronto**
Overcurrent Protection, Note	- Mains - Fuse	

Question:

The note recommends that both supply conductors are fused. How is this recommendation applied?

Decision:

The standard allows the use of a single fuse and therefore, under CB rules this is accepted.

Explanation:

Where a supply is not polarised then equipment employing earthing for protection should be provided with a fuse in both mains conductors.
Evaluators are requested to draw the attention of submittors to this situation and make an informative note in the report as appropriate.