☒ CTL Provisional DECISION SHEET (PDSH)
☐ CTL DECISION SHEET (DSH)

Standard(s) (incl. year)	**Subclause(s)**	**Tracking No.**
OD-5011 (1.1 2019-06)	8.2.1 c .	2178
Category		
General		
Subject	**Keywords**	**Developed by**
Establishing calibration intervals	Initial calibration Statistical analyse	WG1

Question

Can initial calibration intervals for classes of equipment be determined by statistical analysis, where the classes of equipment could be by similar type, manufacturer or model?

Decision

Not in general, only in narrow classes with the same model of an given manufacturer the calibration interval of a new device in the class can be set on the basis of the know statistics of reliability in term for calibration

Explanatory notes

Enquiry: 2178
Date: 2021-04-02

PDSH:
Year of publication: 2025
To be approved at CTL plenary: 2026

DSH
Publication date:

Click here to enter text.

Disclaimer: This document is controlled and has been released electronically.
Only the version on the IECEE Website is the current document version.