CTL Provisional DECISION SHEET (PDSH)

Standard(s) (incl. year)	Subclause(s)	Tracking No.	Publication date
OD 5011	Clause 8.2.1 a)	2246	2024
Category			
General			
Subject	**Keywords**	**Developed by**	**To be approved**
Calibration Interval.	Calibration, timer, interval	WG1	2024 CTL Plenary Meeting

Question

Do digital stopwatches and computer or otherwise electronically based controls for measuring the test duration need to be yearly calibrated

Decision

Since most digital timers and stopwatches are based on a quartz (physically stable), an initial calibration ICO is sufficient for this type of "use". If necessary, an A2 verification with an interval of 24 months can be added together with a battery change.

Explanatory notes

Digital stopwatches and computer or otherwise electronically based controls for measuring the test duration in the environmental simulation do not have to be calibrated, since according to all modern experience they are either "accurate enough" (no significant influence on the test result) or clearly defective. This does not apply to clocks with mechanical elements (movement, hands). They must therefore be calibrated.

The reaction time of the user represents the actually decisive "error size", a tolerance according to CTL OD-5014 with +/- 1% or approx. 1sec/1h is completely sufficient.

Based on DAkkS decision B SK-MB 04/2016.

Disclaimer: This document is controlled and has been released electronically.
Only the version on the IECEE Website is the current document version