

**DUT Model: <blank>**

**DUT Serial Number: <blank>**

**Tested By: testuser on 11/3/25,16:32:36**

**Sequence: TEST\_SEQUENCE Instrument: V73 SN: 39402**

**Passed all tests**

**Step #1 IR\_Test IR @ 1500.0V for 60.0s PASSED**

Leakage: 150.000M $\Omega$  (100.000M $\Omega$  minimum limit)

