



QUORUM
CALIBRATION, LLC

1869 N Hwy 37 Access Rd, George West, TX 78022
361-449-8833 jbarney@qcalibration.net

METER TEST PROVING RECORD

Can / Tank Prover · API MPMS Ch. 4.4

IDENTIFICATION

CUSTOMER	SITE / TERMINAL	ADDRESS	PRODUCT
Global Partners	Chelsea	—	#6 Oil
DATE OF TEST	LAST TEST DATE	THROUGHPUT SINCE	PERFORMED BY
2026-06-08	—	—	Joseph Barney

METER

MAKE	MODEL
Smith	—
SIZE	ID #
—	—
SEAL #	PREVIOUS METER FACTOR
—	—

PROVER & PRODUCT

PROVER SERIAL #	PROVER SIZE
—	1000
PRODUCT GRAVITY (°API)	LAST TOTALIZER
29.3	—
THIS TEST START	THIS TEST FINISH
—	—

RUN DATA

Run	Flow gpm	A Tank gal	C Avg °F	D CTS	E CTL	F Net Prv	G Meter	J Net Mtr	K Err gal	L Err %	M Cu in	N MF
1	—	1000.41	60.0	1.00000	1.00000	1000.41	999.50	999.50	0.91	0.0910	210.2	1.0009

F = A·CTS·CTL · J = G·CTL(invoice) · K = F-J · L = K/J·100 · M = K·231 in³ · MF = F/J · new meter factor = avg MF × previous factor. Open-can atmospheric prover: no pressure correction (CPS/CPL = 1).

RESULTS

PREVIOUS METER FACTOR	AVG METER FACTOR (AS FOUND)	NEW METER FACTOR
—	1.0009	—
AVG ERROR (GAL)	AVG ERROR (CUBIC IN)	REPEATABILITY Δ
0.91	210.2	—

Joseph Barney

PERFORMED BY

WITNESS

Generated 6/8/2026, 5:59:13 PM · Energy Cal · CTS per Shelltable (steel, integer °F, 60 °F = 1.0000); CTL per worksheet coefficient K₀ = 594.5418. Values reproduce the source Excel record.

