pA

5.611
6.362
8.187
8.907
13.809
14.151
16.077
19.795

min

Data File D:\105 F25\Data\chem\2025-11-05 13-27-22 FA25 chem\F-019-72-CL_2.D
Sample Name: CL_2

```
========================================================================
 Acq. Operator   : SYSTEM                      Seq. Line :   19
 Sample Operator : SYSTEM
 Acq. Instrument : Charlie Brown                Location :   72  (F)
 Injection Date  : 11/6/2025 12:29:47 AM            Inj :    9
                                              Inj Volume : 0.5 µl
 Sequence File   : D:\105 F25\Data\chem\2025-11-05 13-27-22 FA25 chem\FA25 chem.S
 Method          : D:\105 F25\Data\chem\2025-11-05 13-27-22 FA25 chem\Whiskey Analysis FRONT
                   only.M (Sequence Method)
 Last changed    : 9/4/2025 3:10:57 PM by SYSTEM
 Method Info     : Whiskey Analysis on DUAL SBP-20 columns. Based on "The Reporter Europe",
                   vol 14, November 2004 International Issue, p  Supelco doc."The Analysis of
                   Alcoholic Beverages on the 30m x 0.25mm ID, 1.0um SPB-20 Capillary Column".
========================================================================
```

```
========================================================================
                          Area Percent Report
========================================================================
 Sorted By            :        Signal
 Multiplier           :        1.0000
 Dilution             :        1.0000
 Do not use Multiplier & Dilution Factor with ISTDs
```

Signal 1: FID1 A,

Peak #	RetTime [min]	Type	Width [min]	Area [pA*s]	Height [pA]	Area %
1	5.611	BV	0.2777	3.59462	1.65277e-1	0.00268
2	6.362	VB	0.3610	3.13328	1.09770e-1	0.00234
3	8.187	BV R	0.3431	1.31953e5	4885.72803	98.54614
4	8.907	VB E	0.0842	59.39811	10.16563	0.04436
5	13.809	BV	0.0832	2.45414	4.53022e-1	0.00183
6	14.151	VB	0.0873	243.69794	44.93510	0.18200
7	16.077	BB	0.0827	1095.48401	203.92223	0.81814
8	19.795	BB	0.0748	538.95111	114.75288	0.40250

```
 Totals :                        1.33899e5   5260.23193
```

```
========================================================================

========================================================================
                          Summed Peaks Report
========================================================================
 Signal 1: FID1 A,
   Empty table.

========================================================================
                       Final Summed Peaks Report
========================================================================
 Signal 1: FID1 A,
```

==
```
Acq. Operator   : SYSTEM                    Seq. Line :  19
Sample Operator : SYSTEM
Acq. Instrument : Charlie Brown             Location :   72  (F)
Injection Date  : 11/6/2025 12:29:47 AM          Inj :   9
                                           Inj Volume : 0.5 µl
Sequence File   : D:\105 F25\Data\chem\2025-11-05 13-27-22 FA25 chem\FA25 chem.S
Method          : D:\105 F25\Data\chem\2025-11-05 13-27-22 FA25 chem\Whiskey Analysis FRONT
                  only.M (Sequence Method)
Last changed    : 9/4/2025 3:10:57 PM by SYSTEM
Method Info     : Whiskey Analysis on DUAL SBP-20 columns. Based on "The Reporter Europe",
                  vol 14, November 2004 International Issue, p  Supelco doc."The Analysis of
                  Alcoholic Beverages on the 30m x 0.25mm ID, 1.0um SPB-20 Capillary Column".
```

==

*** End of Report ***