

# Analysis of Partial Volume Correction on Quantification and Regional Heterogeneity in Cardiac PET

A. Turco, PhD<sup>a</sup>, J. Nuyts, PhD<sup>a</sup>, J. Duchenne, MSc<sup>b</sup>, O. Gheysens, MD, PhD<sup>a,c</sup>, J.U. Voigt, MD, PhD<sup>b,d</sup>, P. Claus, PhD<sup>b</sup>, K. Vunckx, PhD<sup>a</sup>

<sup>a</sup>*KU Leuven - University of Leuven, Department of Imaging and Pathology, Nuclear Medicine and Molecular imaging, Medical Imaging Research Center (MIRC), B-3000 Leuven, Belgium.*

<sup>b</sup>*KU Leuven - University of Leuven, Department of Cardiovascular Sciences, Cardiology, Medical Imaging Research Center (MIRC), B-3000 Leuven, Belgium.*

<sup>c</sup>*University Hospitals Leuven, Department of Nuclear Medicine, B-3000 Leuven, Belgium.*

<sup>d</sup>*University Hospitals Leuven, Department of Cardiovascular Diseases, B-3000 Leuven, Belgium.*

---

## Abstract

*Background.* The partial volume correction (PVC) of cardiac PET datasets using anatomical side information during reconstruction is appealing but not straightforward. Other techniques, which do not make use of additional anatomical information, could be equally effective in improving the reconstructed myocardial activity.

*Methods.* Resolution modelling in combination with different noise suppressing priors was evaluated as a means to perform PVC. Anatomical priors based on a high-resolution CT are compared to non-anatomical, edge-preserving priors (relative difference and total variation prior). The study is conducted on *ex vivo* datasets from ovine hearts. A simulation study additionally clarifies the relationship between prior effectiveness and myocardial wall thickness.

*Results.* Simple resolution modelling during data reconstruction resulted in over- and under- estimation of activity, which hampers the absolute left-ventricular quantification when compared to the ground truth. Both the edge-preserving and the anatomy-based PVC techniques improve the absolute quantification, with comparable results (Student t-test,  $p = 0.17$ ). The

relative tracer distribution was preserved with any reconstruction technique (repeated ANOVA,  $p = 0.98$ ).

*Conclusions.* The use of edge-preserving priors emerged as optimal choice for quantification of tracer uptake in the left-ventricular wall of the available datasets. Anatomical priors visually outperformed edge-preserving priors when the thinnest structures were of interest.

*Keywords:* cardiac PET, system resolution, partial volume correction, quantification

---

## List of abbreviations

PET . . . . . Positron Emission Tomography  
 (HR)CT . . . (High Resolution) Computed Tomography  
 PVE . . . . . Partial Volume Effect  
 PVC . . . . . Partial Volume Correction  
 OSEM . . . . Ordered Subsets Expectation Maximisation (reconstruction algorithm)  
 RR . . . . . Resolution Recovery  
 MAP . . . . . Maximum-a-Posteriori (reconstruction algorithm)  
 TV . . . . . Total Variation (prior)  
 RD . . . . . Relative Differences (prior)

## 1. Introduction

Positron emission tomography (PET) is a sensitive and quantitatively precise preclinical and clinical tool for cardiovascular applications.  $^{18}\text{F}$ -FDG PET is nowadays the most sensitive method to identify myocardial viability [1, 2]. Cardiac PET imaging leads to a panel of clinically relevant parameters that can be interpreted visually (viability, tracer uptake in plaques or sarcoid lesions, regional differences in tracer uptake) but some of these require a quantitative analysis (absolute blood flow, function). Quantification improves reproducibility, reduces inter-observer variability and might enhance the diagnostic accuracy [3].

Despite the wide range of applications for cardiac PET, the resulting images inherently suffer from severe blurring, due to both the finite resolution

of the imaging system (partial volume effect, PVE) and cardiac and respiratory motion during the entire PET examination [4], both of which hamper accurate quantification. The use of partial volume correction (PVC) techniques based on high resolution anatomical information to correct for this PVE has been successfully applied for PET brain imaging and to restore cardiac lesion uptake in a simulation study. In both cases, this resulted in improved quantitative information [5, 6, 7].

The main objective of this study is to evaluate the influence of two types of advanced reconstruction algorithms (anatomy-based and edge-preserving priors) on regional quantification of the left myocardial ventricle (LV) in comparison to the currently used clinical algorithms (with or without resolution modelling). All algorithms were applied on *ex vivo* datasets excised from sheep with asymmetric cardiac remodelling. The use of *ex vivo* (static) datasets allowed to rule out quantification errors caused by cardiac or respiratory motion, and could therefore be considered as an ideal version of a single frame from a doubly gated *in vivo* study.

The animal model used in this work was an ovine model of dilated cardiomyopathy with asymmetric left ventricular remodelling due to a left-bundle branch block-like conduction delay. Some of the typical features of this pathology are a dilation of the LV chamber over time, together with a thinning of the septal wall and a thickening of the lateral wall of the LV during the progression of the disease. Such variations in the wall thicknesses were hypothesised to generate problems with relative tracer uptake quantification, due to the partial volume effects which are expected to mostly affect the thinned walls. The application of PVC techniques in hearts which underwent such kind of remodelling should best demonstrate the recovery of activity in the thinnest walls, provided that the final wall thinning is enough to cause severe PVEs.

## 2. Methods

### 2.1. Animal model and experimental design

We used an ovine model of rapid pacing-induced dilated cardiomyopathy with left-bundle branch block-like morphology leading to asymmetric remodelling of the septal and lateral wall. All experiments and procedures were approved by the local ethical committee.

Each of the thirteen available sheep was sacrificed and their heart was excised. The excised, *ex vivo* hearts were filled with a non-attenuating, hard-

ening poly-urethane foam in order to ensure a stable shape of the heart in time. Each heart was filled once throughout the experiment, hence the heart shape was always the same throughout the *ex vivo* scans of the same animal. A pilot *ex vivo* experiment had been previously carried out to confirm that the foam would not keep on expanding throughout the duration of the imaging process. The following *ex vivo* scans were performed:

- a 15-minute acquisition on a small-animal PET scanner (Siemens Focus220 microPET), which served as gold standard because of the high spatial resolution of this PET system (1.5 mm), resulting in images that are almost free from resolution-related blurring. A transmission scan with a rotating  $^{57}\text{Co}$  point source was also obtained for scatter correction.
- a 20-minute PET scan on the clinical PET/CT scanner (Siemens Biograph 16, Hirez). These datasets were reconstructed with different algorithms and compared to the ground truth images provided by the microPET.
- a high-dose, high-resolution CT (HRCT) on the clinical PET/CT scanner. It was used for accurate attenuation correction and for PVC of both the clinical and the microPET datasets. In the *ex vivo* hearts the blood is removed from the ventricular cavities and the polyurethane foam is injected in their place. This allows a clear distinction of the myocardium, thus ensuring an efficient PVC. No contrast-enhanced CT is therefore needed in this case.

## 2.2. Simulation study

A simulation study was designed to clarify whether there is a relationship between the apparent thickness of the reconstructed structures and the different reconstruction algorithms.

For this purpose, several cylindrical phantoms resembling the shape of the left ventricle were designed, all with a height of 65 mm and an outer diameter of 30 mm. The cylindrical walls of each phantom had a uniform thickness, which was different for each phantom and ranged between 1 and 12 mm (representative of the thicknesses that can be encountered in a cardiac setting). Figure 1 shows one of the designed phantoms.

The walls of the phantom were filled with a known and fixed activity concentration of 4  $kBq/cc$ . A second identical phantom was then created




Figure 1: Transaxial (left) and axial (right) view of one of the phantoms simulated to demonstrate the effect of the reconstruction algorithms against different phantom thicknesses. The arrow with *Variable thickness* indicates the part of the phantom that was modified in the subsequent simulations to study the effect of different wall thicknesses (in this example, the wall thickness is 8 mm).

to simulate a uniform attenuation map and filled with a known and fixed attenuation value of 0.0999428 (attenuation  $\times \text{cm}^{-1}$ ), which corresponds to the mean attenuation value found for muscle at 511 keV [8]. A third copy of the phantom was used to simulate the anatomical prior information, perfectly aligned to the emission dataset.

A simulated acquisition of each of these numeric phantoms was performed, modelling the physical phenomena of limited spatial resolution, attenuation and detector normalization. Scatter and randoms were not modelled to exclude their influence from the analysis. A noise-free sinogram was obtained for each phantom. The resulting total sinogram counts are between 0.5 and 2.5 million counts, depending on the phantom thickness. This is in the same order of magnitude as the total sinogram counts of the available heart datasets with the lowest uptake and a 20 min data acquisition. One noisy sinogram was generated for each thickness-phantom by adding Poisson noise to the noise-free sinogram.

### 2.3. Image reconstruction

#### *Hirez PET/CT datasets*

Both simulated and real measured datasets were reconstructed using the same iterative reconstruction scheme ( $[3,2,2]$  iterations  $\times$   $[42,24,1]$  subsets). A realistic, in-house developed model of the Hirez PET system was used both for the simulated acquisition of the numeric phantoms and for the reconstruction of all the Hirez PET datasets. When requested, the spatial resolution of the scanner was modelled as a Gaussian function whose full-width at half-maximum corresponds to the spatial resolution of the scanner (4.3 mm and

4.5 mm in the transaxial and axial direction, respectively). The voxel size of the reconstructed PET images was set to  $1.35 \times 1.35 \times 1.35 \text{ mm}^3$  for all datasets.

An overview of the different algorithms used to reconstruct each of the available sinograms is listed below:

1. iterative reconstruction (ordered subsets expectation maximisation, **OSEM**) without resolution recovery (RR)
2. OSEM with RR (**OSEM + RR** in the rest of this work), which takes into account the resolution of the scanner during the PET reconstruction.
3. iterative reconstruction with prior information (maximum-a-posteriori, MAP) with RR, with CT-based side anatomical information. The anatomical information was incorporated in the reconstruction process by means of the asymmetrical **Bowsher** prior [9, 10]. The Bowsher prior is parametrised by two values, namely the weight (w) and the number of neighbours (n), which define the strength of the prior and the size of its smoothing mask respectively.
4. MAP with RR in combination with a non-anatomy based prior (known as *relative differences* (**RD**) prior [11]), parametrised by a weight (w) and an edge-preserving parameter ( $\gamma$ ).
5. MAP with RR, in combination with a non-anatomy based prior (known as *isotropic total variation prior* (**TV**) [12]), parametrised by a weight (w).

A range of prior parameter values was chosen based on a simulation study, where reconstructions with a similar count level were compared to a ground truth dataset in terms of bias and noise. A visual inspection of the datasets confirmed that the chosen range was an appropriate compromise between noise reduction and excessive smoothing of the structures.

#### *microPET datasets*

The microPET list-mode datasets of the *ex vivo* hearts were reconstructed using a MAP reconstruction algorithm (5 iterations  $\times$  28 subsets) with resolution recovery [13] and asymmetrical Bowsher prior (w = 50, n = 9). We chose to apply a prior because resolution recovery alone is known to reduce the apparent thickness of the structures and to possibly hamper quantification [14]. Moreover, Bowsher reconstructions have been shown to improve the recovery of activity concentrations in previous studies, e.g. in [6]. In order to

144 corroborate such findings and choose which microPET reconstruction to use  
 145 as ground truth, we performed a preliminary simulation study. The Bowsher  
 146 reconstruction with the aforementioned parameters was found to minimise  
 147 the bias and the noise when compared to a simulated ground truth. It was  
 148 therefore chosen as the reconstruction method for the microPET datasets in  
 149 this study.

150 Attenuation correction of the microPET was based on the HRCT and  
 151 performed, together with scatter and sensitivity correction, during recon-  
 152 struction. The data were also precorrected for randoms and deadtime. The  
 153 microPET images were reconstructed with a voxel size of  $0.4745 \times 0.4745 \times$   
 154  $0.796 \text{ mm}^3$ . For comparison to the Hirez datasets, they were rigidly aligned  
 155 and resampled to match the Hirez.

#### 156 2.4. Image post-processing

157 A polar map (or bull’s eye) representation of the simulated and *ex vivo*  
 158 reconstructions was chosen to illustrate the regional and absolute differences  
 159 within the myocardium. To create the bull’s eye plots, the software must  
 160 compute a single value for each longitudinal and angular position within the  
 161 LV. This value must represent the activity over the LV wall thickness at each  
 162 of those positions. To do so, the LV wall portion enclosed by the endocardial  
 163 and the epicardial contours is considered for each angular and longitudinal  
 164 position. The definition of the endocardial and epicardial contours is done  
 165 automatically using a model-based approach, as described in [15]. The maxi-  
 166 mum activity value or the mean activity value along each of those portions is  
 167 computed (see [7] for further details on the process). This value is then used  
 168 to represent the activity value over the myocardial thickness. Both types of  
 169 polar maps were created in this work, by taking the maximum or the mean  
 170 count over the LV thickness (*max-count* and *mean-count* polar maps in the  
 171 rest of this work, respectively).

#### 172 2.5. Image analysis

##### 173 Measured *ex vivo* datasets

174 The **absolute** values of the *ex vivo* reconstructions were compared. The  
 175 mean difference between the microPET polar map and the *ex vivo* Hirez  
 176 polar maps, relative to the mean activity of the microPET polar map, was  
 177 computed and plotted, for each animal. The values extracted from the mean-  
 178 count polar maps were compared to the values extracted from the max-count  
 179 polar maps.

180 For every polar map, the S/L ratio was additionally calculated by divid-  
 181 ing the mean activity in the free septal wall (segments 8, 9 in a conventional  
 182 17-segment polar map [16]) by the mean activity in the free lateral wall (seg-  
 183 ments 11, 12). This analysis aimed to quantify whether the prior-based re-  
 184 constructions improve or preserve the **regional variations** of activity within  
 185 the heart, when compared to the S/L ratio computed on the polar map of  
 186 the corresponding ground truth reconstructions.

187 Although the analysis of the septal vs lateral walls is commonly performed  
 188 to describe regional inhomogeneity in the heart function and metabolism [17],  
 189 a more general overview of the inhomogeneities was also provided. To this  
 190 end, each of the available *ex vivo* max-count polar maps was considered, and  
 191 divided by its maximum to obtain a *normalized* polar map (unit-less, ranging  
 192 between 0 and 1). For each sheep, all the normalized polar maps from the  
 193 Hirez scanner, obtained with the different reconstruction techniques, were  
 194 compared to the normalized polar map of the ground truth. The resulting  
 195 percentage bias was computed, per segment. The mean of the segmental bi-  
 196 ases was computed over all the sheep, together with the segmental standard  
 197 deviation of the bias over all sheep. Two corresponding polar maps were  
 198 generated for each reconstruction algorithm, whose segmental values repre-  
 199 sent the mean bias and the mean standard deviation of the normalized polar  
 200 maps of the Hirez against the ground truth.

## 201 *Simulated datasets*

202 The mean activity in the middle ring of each max-count polar map of  
 203 the simulated phantoms was plotted for the different phantom thicknesses in  
 204 comparison to the activity expected in that region ( $4 \text{ kBq/cc}$ ).

205 Additionally, profiles through the wall were taken in the cylindrical part  
 206 of the heart phantom (Figure 1). An average of the 20 central subsequent  
 207 axial profiles was considered, to minimise the contribution of noise to the  
 208 profiles. These mean profiles were plotted for all different reconstruction  
 209 algorithms.

## 210 *2.6. Statistical methods*

211 A repeated ANOVA between the different reconstructions of the mea-  
 212 sured datasets was performed, to establish whether a statistical difference in  
 213 the mean absolute activity differences could be found, and the resulting p  
 214 and F values were reported (significance level: 0.05). An additional Student  
 215 t-test with posthoc Bonferroni correction helped to point out the algorithms






Figure 2: Representative images of an *ex vivo* sheep heart (sheep4238) reconstructed with different algorithms. On the right, the corresponding max-count polar maps for the same sheep are shown. The red box highlights the ground truth dataset (microPET). Color scale in Bq/cc. OSEM: ordered-subsets expectation-maximization (iterative reconstruction) algorithm, RR: resolution recovery, RD: relative differences prior, TV: total variation prior.

216 which significantly differed from the state-of-the-art reconstruction technique  
 217 (OSEM + RR). A repeated ANOVA test between the mean S/L ratios com-  
 218 puted from the various reconstructions was also performed, and the resulting  
 219 p and F values were reported (significance level: 0.05).

### 220 3. Results

#### 221 *Measured ex vivo datasets*

222 A representative dataset with a selection of the performed reconstructions  
 223 is shown in Figure 2. The use of well tuned edge-preserving priors (RD, TV)  
 224 seems to work as well as a (well tuned) Bowsher prior, even though the  
 225 Bowsher reconstructions show moderately sharper edges. In all cases, the  
 226 visual assessment of the different reconstructions did not reveal dramatic  
 227 changes in the relative activity distribution. The polar maps obtained for all  
 228 the ex vivo datasets were also compared visually (right column of Figure 2)  
 229 and confirmed the previous observations.

230 The accurate **absolute recovery** of the activity concentrations within  
 231 the LV was then considered (Figure 3), which confirms an overall underes-  
 232 timation of the mean LV activity when no resolution recovery is modelled  
 233 and a significant mean overestimation of the LV activity for the reconstruc-  
 234 tions with resolution recovery alone (repeated ANOVA: F-value=150, p-value  
 235  $< 0.001$  (significant)). The T-test between TV and the latter (OSEM+RR)  
 236 gives a p-value  $< 0.001$  (significant), whereas the same test between TV and  
 237 Bowsher  $w=30, n=9$  has a p-value of 0.17 (not significant). This ranking  
 238 holds true when the mean-count polar maps are used (bottom pane of Fig-  
 239 ure 3), but the differences between the different algorithms are more subtle.  
 240 However, we can clearly appreciate the stabilizing role of the regularization  
 241 included during the reconstruction process.

242 The **S/L ratios** were further calculated for the polar maps of each of  
 243 the clinical reconstructions. They were compared to the ground truth S/L  
 244 ratio extracted from the corresponding microPET reconstruction. Within  
 245 each animal, the S/L ratios do not significantly change with different recon-  
 246 struction algorithms, further supporting our visual evaluation of the polar  
 247 maps. The ANOVA test showed no significant differences in the mean S/L  
 248 ratios recovered from the different reconstruction algorithms ( $F = 0.25$ ,  $p =$   
 249  $0.98$ ).

250 The polar maps describing the mean bias and the mean standard de-  
 251 viation of the bias, over all sheep, are reported in Figure 4 for 4 selected




Figure 3: Mean difference between the polar maps of the ground truth (microPET) and the polar maps of the *ex vivo* Hirez images, reconstructed with the various algorithms, relative to the mean activity of the polar map of the corresponding microPET. Connected dots of the same color represent individual animals. The red solid bar is the mean value over all animals for each of the considered algorithms. The use of a different way to obtain the polar maps (max-count (top) or mean-count (bottom) polar maps) changes the amount of error that is computed, especially for OSEM+RR (red arrows). The use of edge-preserving or anatomical priors leads to a more accurate and more stable absolute quantification of the activity in the LV, regardless of the type of polar map that is used. The red boxes highlight this behaviour for two representative cases.

252 algorithms: Bowsher (weight=30, n=9), TV (weight=0.005) and for the it-  
 253 erative reconstruction with and without resolution recovery. Most of the  
 254 individual biases are within the  $\pm 0.1$  threshold, thus indicating very small  
 255 differences (up to 10%) in the reconstructed LV homogeneity.




Figure 4: Mean bias (left side of each quadrant) and mean standard deviation of the bias (right side in each quadrant) over all animals, extracted from the normalized max-count polar maps of the reconstructions using a selection of the different algorithms. OSEM: iterative reconstruction, with or without resolution recovery (RR); TV: non-anatomical prior, Bowsher: anatomical prior. The bias is a fraction of the maximum of the normalized polar maps, i.e. a fraction of 1.

### 256 *Simulated datasets*

257 The analysis of the mean reconstructed activity at different phantom  
 258 thicknesses is in Figure 5. As hypothesized, the thinnest phantoms benefit  
 259 more from the additional use of anatomical information (black line), while  
 260 the non-anatomical priors – or the anatomical priors where the number of  
 261 neighbours over which the smoothing is performed is too large – blur over the  
 262 edges and therefore struggle to fully restore the activity within the walls of  
 263 the thinnest phantoms. The iterative reconstruction with resolution recovery

264 alone displays the typical overestimation of activity at a thickness of 6-8 mm  
 265 (roughly double the spatial resolution of the system), which becomes less  
 266 apparent with increasing wall thickness of the cardiac phantom ( $\geq 10$  mm).




Figure 5: Profiles over one wall of the heart phantom (simulated homogeneous activity of 4  $kBq/cc$ ), for a selection of phantom thicknesses. To emphasize the effect of the parameter choice in the Bowsher reconstructions, especially at small thicknesses, a high weight and two very different numbers of neighbours ( $n$ ) are shown.

267 The analysis of the mean recovered value in the max-count bull’s eye plots  
 268 of the heart phantom at different thicknesses (Figure 6) confirms the findings  
 269 of the profile analysis and of the analysis of the *ex vivo* measured datasets.

270 The visual inspection of the thinnest structures of the sheep datasets (e.g.  
 271 right ventricle) confirms the improved reconstruction of such areas when an  
 272 adequately chosen anatomical prior is used. Figure 7 shows a comparison  
 273 between the short-axis slices of two Bowsher and two TV reconstructions.

## 274 4. Discussion

275 The use of anatomical side information for the improved quantification  
 276 of cardiac PET datasets stemmed from previous, successful experiences in  
 277 brain PET imaging, where similar techniques have been applied during the  
 278 reconstruction of PET datasets [6]. The results obtained in the current study  
 279 are in line with previous findings and add interesting insights into the dis-  
 280 criminating factors for choosing an anatomy-based prior.

281 The modelling of the resolution during reconstruction is essential for deal-  
 282 ing with the PVE. However, our findings confirm that the resolution recovery




Figure 6: Mean value in the max-count polar maps of the heart phantoms, at increasing phantom thicknesses. The Bowsher algorithm, with a well chosen neighborhood, is the one that best improves the recovery of the mean activity concentration starting from a thickness of 2 mm.

is an ill-posed problem [18] and, alone, is not sufficient for accurate image quantification. Some kind of regularization is additionally needed to produce more quantitatively accurate results. In this study, the use of priors proves to improve the absolute quantification of the left ventricles, when compared to the state-of-the-art iterative reconstruction algorithms (with or without resolution recovery). This paper highlights the role of both anatomical (Bowsher) and non-anatomical edge-preserving priors (RD, TV) applied during the reconstruction of the datasets in stabilizing the resulting images. Even when the weights or the number of neighbours chosen are not fully optimized for the structures of interest (e.g. too many neighbours, or too bland prior weight), the error calculated in the enhanced reconstructions is milder than the one obtained in the reconstructions where no priors are used.

For regional analysis and relative quantification of the activity distribution within the left ventricle, on the other hand, the use of either reconstruction algorithm (with or without prior information) was sufficient and provided similar information. The S/L ratios are equally well resolved when



Figure 7: Comparison between non-anatomical (TV) and CT-based Bowsher reconstructions, for three different sheep datasets (from left to right column: sheep4238, sheep1869 and sheep1132). For both algorithms, the same short-axis slice and the same maximum is shown (they differ for the different animals, to maximize visibility).

299 compared to the ground truth ratio, no matter the reconstruction algorithm  
 300 that is chosen for reconstructing the datasets obtained on the clinical scanner.  
 301 It is worth mentioning that the Bowsher priors presented in the study on  
 302 the wall thickness for demonstrative purposes use a considerably high weight  
 303 to achieve such results. In a more realistic scenario where these techniques  
 304 would be applied, i.e. a motion-corrected acquisition with the use of anatom-  
 305 ical information obtained from a dedicated scanner, the use of a high weight  
 306 would be risky and could cause artefacts if the alignment between the PET  
 307 and the anatomical dataset is not accurate. To present, the prior param-  
 308 eters need to be tuned in a case-by-case fashion to obtain optimal results,  
 309 which can be a limiting factor for a realistic clinical workflow. A way to  
 310 automatically select the best prior parameters was out of the scope of this  
 311 work and should be subject of further investigation. Additionally, it is nec-  
 312 essary to underline the importance of a correct image alignment and motion  
 313 correction, before proceeding to anatomy-based PET reconstruction. With  
 314 these regards, truly simultaneous PET/MR systems might represent a great  
 315 improvement in the direction of accurate alignment, provided that the tech-  
 316 nical advances in MRI imaging will reduce the geometrical distortions and  
 317 artefacts that nowadays still affect the gated MR measurements [19].

318 A limitation of this study is the sole use of *ex vivo*, static cardiac datasets  
 319 for the assessment of the usefulness of partial volume correction techniques.  
 320 In patients, the motion of the heart due to the beating and the breathing  
 321 during the scan can create additional blurring of the PET datasets and needs  
 322 to be taken into account and fully corrected if accurate quantification is  
 323 aimed at. We expect a worsening of the quantification performances of any  
 324 of the presented algorithms in the case where the correction of the motion is  
 325 inaccurate or missing. The use of *in vivo* datasets, which we acquired within  
 326 the same study as the *ex vivo* datasets, should help to shed further light into  
 327 the applicability of the presented techniques to images of the beating heart.  
 328 The analysis of the *in vivo* datasets is ongoing. Particularly, given that no  
 329 partial volume correction technique should be applied to datasets that are  
 330 not corrected for motion, we are investigating on the applicability of partial  
 331 volume correction techniques in combination with dual-gating techniques.  
 332 The results of the *in vivo* study will be made available upon completion.  
 333 A particularly important conclusion from the present *ex vivo* study, where  
 334 no motion is present and alignment with anatomical data is perfect, is that  
 335 non-anatomical priors and anatomical, CT-based priors perform equally well  
 336 for the quantification of LV activity. As the current *ex vivo* experiment can  
 337 be considered as an ideal dual-gated *in vivo* experiment – where the scatter,  
 338 attenuation and motion correction have been perfectly accounted for –, we  
 339 believe that the application of non-anatomy based priors is a good starting  
 340 point for truly dual-gated *in vivo* studies where the LV walls are of interest.  
 341 This is advantageous as it simplifies the overall acquisition and reconstruction  
 342 process.

343 Finally, although the application of non-anatomical priors yielded accept-  
 344 able quantification results on the available animal cohort (i.e. hearts without  
 345 lesions and with sufficiently thick LV walls), the application of anatomical  
 346 information for cardiac PET imaging is not to be fully excluded, particularly  
 347 when non-transmural lesions are present in the considered datasets [7] and  
 348 in general when the thinnest structures are of interest.

## 349 5. Conclusions

350 We have demonstrated the necessity of prior information during the iter-  
 351 ative reconstruction of *ex vivo* cardiac datasets, to be performed in combina-  
 352 tion with the modelling of the scanner’s resolution. The use of side anatom-  
 353 ical information was not essential for our animal datasets, and a well-tuned



354 non-anatomical prior was sufficient for the recovery of the LV activity.

355 The analysis of the simulated data demonstrates that the Bowsher prior  
356 could have potential use in the recovery, delineation and analysis of the PET  
357 datasets, particularly if the thickness of the considered structures is expected  
358 to drop below the 4 mm.

## 359 **6. New knowledge gained**

360 This work addresses the previously unanswered question regarding the  
361 usefulness of anatomical and non-anatomical priors for the quantification of  
362 cardiac PET images. It provides guidelines for the reconstruction of cardiac  
363 emission datasets, partially confirming the results of previous literature on  
364 brain PVC, partially adding new insights into the usefulness of the prior  
365 information in the reconstruction of different cardiac structures. The study  
366 design using *ex vivo* cardiac datasets acquired on both a pre-clinical and a  
367 clinical scanner is innovative and original.

## 368 **7. Competing interests**

369 The authors declare that they have no competing interests.

## 370 **8. Author's contributions**

371 AT was responsible for the study design, the simulation setup, the recon-  
372 structions and the data collection and analysis, and drafted the manuscript.  
373 JN and KV assisted with the study design, the analysis of data and the  
374 careful revision of the manuscript. JUV, PC, JD and OG participated in  
375 the study design and critically revised the manuscript. All authors read and  
376 approved the final manuscript.

## 377 **9. Acknowledgements**

378 AT is a PhD fellow of the Research Foundation - Flanders (FWO). OG  
379 (1831812N) and JUV are senior clinical investigators of the Research Founda-  
380 tion - Flanders. KV is a post-doctoral researcher of the Research Foundation  
381 - Flanders.

382 The authors wish to thank Charles Watson, Judd Jones et al. for their  
383 help with the Siemens data processing, and Filip Rega for the development  
384 of the animal model used in this work.

- 385 [1] A. F. Schinkel, J. J. Bax, D. Poldermans, A. Elhendy, R. Ferrari, S. H.  
386 Rahimtoola, Hibernating myocardium: diagnosis and patient outcomes,  
387 *Current problems in cardiology* 32 (7) (2007) 375–410.
- 388 [2] J. Knuuti, H. R. Schelbert, J. J. Bax, The need for standardisation of  
389 cardiac FDG PET imaging in the evaluation of myocardial viability in  
390 patients with chronic ischaemic left ventricular dysfunction, *European*  
391 *journal of nuclear medicine and molecular imaging* 29 (9) (2002) 1257–  
392 1266.
- 393 [3] G. El Fakhri, A. Kardan, A. Sitek, S. Dorbala, N. Abi-Hatem, Y. La-  
394 houd, A. Fischman, M. Coughlan, T. Yasuda, M. F. Di Carli, Repro-  
395 ducibility and accuracy of quantitative myocardial blood flow assess-  
396 ment with  $^{82}\text{Rb}$  PET: comparison with  $^{13}\text{N}$ -ammonia PET, *Journal of*  
397 *Nuclear Medicine* 50 (7) (2009) 1062–1071.
- 398 [4] M. Soret, S. L. Bacharach, I. Buvat, Partial-volume effect in PET tumor  
399 imaging, *J. Nucl. Med.* 48 (6) (2007) 932–945.
- 400 [5] A. M. Alessio, P. E. Kinahan, Improved quantitation for PET/CT image  
401 reconstruction with system modeling and anatomical priors., *Med Phys*  
402 33 (2006) 4095:4103.
- 403 [6] K. Vunckx, A. Atre, K. Baete, A. Reilhac, C. M. Deroose, K. Van Laere,  
404 J. Nuyts, Evaluation of three MRI-based anatomical priors for quanti-  
405 tative PET brain imaging., *IEEE Trans Med Imaging* 31 (3) (2012)  
406 599612.
- 407 [7] A. Turco, J. Nuyts, O. Gheysens, J. Duchenne, J.-U. Voigt, P. Claus,  
408 K. Vunckx, Lesion quantification and detection in myocardial  $^{18}\text{F}$ -FDG  
409 PET using edge-preserving priors and anatomical information from CT  
410 and MRI: a simulation study, *EJNMMI physics* 3 (1) (2016) 1–32.
- 411 [8] J. H. Hubbell, S. M. Seltzer, N. I. of Standards, T. (U.S.), Tables  
412 of X-ray mass attenuation coefficients and mass energy-absorption  
413 coefficients 1 keV to 20 MeV for elements  $Z=1$  to 92 and 48  
414 additional substances of dosimetric interest, U.S. Department of  
415 Commerce, Technology Administration, National Institute of Stan-  
416 dards and Technology Gaithersburg, MD, web version. edn., URL

- 417 <http://physics.nist.gov/PhysRefData/XrayMassCoef/cover.html>,  
418 1996.
- 419 [9] J. Bowsher, H. Yuan, L. Hedlund, T. Turkington, G. Akabani, A. Badea,  
420 W. Kurylo, C. Wheeler, G. Cofer, M. Dewhirst, G. Johnson, Utilizing  
421 MRI information to estimate F18-FDG distributions in rat flank tumors,  
422 in: Nuclear Science Symposium Conference Record, 2004 IEEE, vol. 4,  
423 2488–2492 Vol. 4, 2004.
- 424 [10] K. Vunckx, J. Nuyts, Heuristic modification of an anatomical Markov  
425 prior improves its performance, in: Nuclear Science Symposium Confer-  
426 ence Record (NSS/MIC), 2010 IEEE, 3262–3266, 2010.
- 427 [11] J. Nuyts, D. Beque, P. Dupont, L. Mortelmans, A concave prior pe-  
428 nalizing relative differences for maximum-a-posteriori reconstruction in  
429 emission tomography, Nuclear Science, IEEE Transactions on 49 (1)  
430 (2002) 56–60.
- 431 [12] A. Chambolle, T. Pock, A First-Order Primal-Dual Algorithm for Con-  
432 vex Problems with Applications to Imaging, Journal of Mathematical  
433 Imaging and Vision 40 (1) (2011) 120–145.
- 434 [13] Y.-C. Tai, A. Ruangma, D. Rowland, S. Siegel, D. F. Newport, P. L.  
435 Chow, R. Laforest, Performance evaluation of the microPET focus: a  
436 third-generation microPET scanner dedicated to animal imaging, Jour-  
437 nal of nuclear medicine 46 (3) (2005) 455–463.
- 438 [14] A. M. Alessio, A. Rahmim, Resolution modeling enhances PET imaging,  
439 Med. Phys 40 (2013) 120601.
- 440 [15] J. Nuyts, L. Mortelmans, P. Suetens, A. Oosterlinck, M. de Rou, Model-  
441 based quantification of myocardial perfusion images from SPECT, J.  
442 Nucl. Med. 30 (12) (1989) 1992–2001.
- 443 [16] M. D. Cerqueira, N. J. Weissman, V. Dilsizian, A. K. Jacobs, S. Kaul,  
444 W. K. Laskey, D. J. Pennell, J. A. Rumberger, T. Ryan, M. S. Verani,  
445 et al., Standardized myocardial segmentation and nomenclature for to-  
446 mographic imaging of the heart a statement for healthcare professionals  
447 from the cardiac imaging committee of the Council on Clinical Cardi-  
448 ology of the American Heart Association, Circulation 105 (4) (2002)  
449 539–542.

- 450 [17] K. Russell, M. Eriksen, L. Aaberge, N. Wilhelmsen, H. Skulstad, E. W.  
 451 Remme, K. H. Haugaa, A. Opdahl, J. G. Fjeld, O. Gjesdal, T. Ed-  
 452 vardsen, O. A. Smiseth, A novel clinical method for quantification of  
 453 regional left ventricular pressure–strain loop area: a non-invasive index  
 454 of myocardial work, *European Heart Journal* ISSN 0195-668X.
- 455 [18] J. Nuyts, Unconstrained image reconstruction with resolution modelling  
 456 does not have a unique solution, *EJNMMI physics* 1 (1) (2014) 1.
- 457 [19] S. G. Nekolla, C. Rischpler, K. P. Kunze, PET/MRI for Cardiac Imag-  
 458 ing: Technical Considerations and Potential Applications, in: *Molecular*  
 459 *and Multimodality Imaging in Cardiovascular Disease*, Springer, 29–48,  
 460 2015.