

Regional accuracy of ZTE-based attenuation correction in static [^{18}F]FDG and dynamic [^{18}F]PE2I brain PET/MR

G. Schramm^{1,*}, M. Koole¹, S. M. A. Willekens¹, A. Rezaei¹,
D. Van Weehaeghe¹, G. Delso², R. Peeters³, N. Mertens¹, J. Nuyts¹, and
K. Van Laere¹,

¹KU/UZ Leuven, Department of Imaging and Pathology, Division of Nuclear Medicine, Leuven, Belgium

²GE Healthcare, Cambridge, UK

³UZ Leuven, Division of Radiology, Leuven, Belgium

Correspondence*:

Georg Schramm

georg.schramm@kuleuven.be

ABSTRACT

3

Accurate MR-based attenuation correction (MRAC) is essential for quantitative PET/MR imaging of the brain. In this study, we analyze the regional bias caused by MRAC based on Zero-Echo-Time MR images (ZTEAC) compared to CT-based AC (CTAC) static and dynamic PET imaging. In addition the results are compared to the performance of the current default Atlas-based AC (AtlasAC) implemented in the GE SIGNA PET/MR.

Methods: Thirty static [^{18}F]FDG and eleven dynamic [^{18}F]PE2I acquisitions from a GE SIGNA PET/MR were reconstructed using ZTEAC (using a research tool, GE Healthcare), single-subject AtlasAC (the current default AC in GE's SIGNA PET/MR) and CTAC (from a PET/CT acquisition of the same day). In the 30 static [^{18}F]FDG reconstructions, the bias caused by ZTEAC and AtlasAC in the mean uptake of 85 anatomical volumes of interest (VOIs) of the Hammers' atlas was analyzed in PMOD. For the eleven dynamic [^{18}F]PE2I reconstructions, the bias caused by ZTEAC and AtlasAC in the non displaceable binding potential BP_{nd} in the striatum was calculated with cerebellum as the reference region and a simplified reference tissue model.

Results: The regional bias caused by ZTEAC in the static [^{18}F]FDG reconstructions ranged from -8.0% to +7.7% (mean 0.1%, SD 2.0%). For AtlasAC this bias ranged from -31.6% to +16.6% (mean -0.4%, SD 4.3%). The bias caused by AtlasAC showed a clear gradient in the cranio-caudal direction (-4.2% in the cerebellum, +6.6% in the left superior frontal gyrus). The bias in the striatal BP_{nd} for the [^{18}F]PE2I reconstructions ranged from -0.8% to +4.8% (mean 1.5%, SD 1.4%) using ZTEAC and from -0.6% to +9.4% using AtlasAC (mean 4.2%, SD 2.6%).

Conclusion: ZTEAC provides excellent quantitative accuracy for static and dynamic brain PET/MR, comparable to CTAC, and is clearly superior to the default AtlasAC currently implemented in the GE SIGNA PET/MR.

Keywords: MR-based attenuation correction, PET/MR, PET quantification, PET reconstruction, Molecular Imaging

1 INTRODUCTION

Since the introduction of combined PET/MR, accurate attenuation correction (AC) for brain imaging has always been a field of active research. Neglecting higher bone attenuation of the skull in the first generation segmentation-based AC methods used in product implementations led to a substantial spatially-varying bias in the reconstructed tracer uptake [1]. To include patient-specific information about higher bone attenuation, two concepts for MR-based attenuation correction (MRAC) were investigated by different research groups. On the one hand, ultra short echo time (UTE) MR sequences that allow to generate signal in cortical bone were used to segment bone structures in the skull [2, 3, 4, 5, 6, 7]. On the other hand, the use of single [8, 9, 10] or multi MR-CT atlas [11, 12] information to generate attenuation images including higher bone attenuation were proposed. Recently, Ladefoged et al. [13] showed in a multi-center evaluation that the bias introduced by MRAC in brain PET/MR imaging can be reduced to $\pm 5\%$ when using different second generation atlas- or UTE-based AC techniques developed by different research groups.

Weiger et al. [14] and Wiesinger et al. [15] showed that zero echo time (ZTE) MR sequences have great potential in imaging materials with short T_2^* such as cortical bone. Since ZTE sequences only use a single echo, their acquisition time is substantially shorter compared to UTE sequences that usually acquire two echos. In addition, faster switching from transmit to receive in the ZTE sequence minimizes loss of signal in tissues with short T_2^* relaxation times such as cortical bone. Due to the use of minimal gradient switching, ZTE is less prone to eddy current artifacts than UTE [15, 16]. Moreover, a correlation between the ZTE MR signal intensity and CT Hounsfield units (HU) in cortical bone was demonstrated in [15].

Consequently, ZTE MR imaging is very promising for accurate AC in brain PET/MR. Delso et al. [17] showed that ZTE-based skull segmentation, which is needed to generate attenuation images including higher bone attenuation, is feasible. Boydev et al. [18] showed that the use of ZTE MR images in their atlas-based prediction of pseudo CTs improved the correctness of the pseudo CTs for radiation therapy planning in case of bone resection surgery prior to the radiation therapy compared to using T1-weighted MR images as input.

Moreover, Sekine et al. [16], Khalife et al. [19], Yang et al. [20], Leynes et al. [21], Wiesinger et al. [22] recently demonstrated that the quantitative accuracy of PET images reconstructed with ZTE-based attenuation images is high. All groups investigated pilot studies with small patient cohorts (10, 16, 12, 6, and 5 subjects, respectively) and evaluated static FDG PET images.

So far, no evaluation of ZTE-AC for absolute quantification of dynamic receptor studies (e.g. in terms of non-displaceable binding potential or distribution volume) has been published. The influence of attenuation correction on parameters derived from kinetic modeling is more complex especially in case when reference tissue models are used. In those cases, it is important to have accurate attenuation correction for the target region (e.g. the striatum) as well as for the reference region (e.g. the cerebellum).

To study the influence of ZTE-based AC on the accuracy of tracer kinetic modeling using the simplified reference tissue model, we analyzed eleven dynamic PET/MR acquisitions with the highly selective dopamine transporter tracer [^{18}F]PE2I. In addition, we investigated the regional quantitative accuracy of ZTE-based AC in 30 static [^{18}F]FDG PET/MR acquisitions.

2 MATERIALS AND METHODS

2.1 Subjects

We included 48 subjects that participated in two ongoing PET/MR research protocols in the context of neurodegenerative diseases. Thirty-four patients suspected for dementia were investigated with a static [^{18}F]FDG PET/MR protocol between October 2016 and June 2017. Three patients were excluded from this comparison study due to dental implants which led to metal artifacts in the MR images. In one case the patient was positioned too low in the head coil which led to very low ZTE MR signal in the caudal end of the head due to low coil sensitivity in that region. This case was excluded as well. The mean age of the remaining 30 patients was 63 y (range 40–77 y). In addition, we analyzed 14 dynamic [^{18}F]PE2I acquisitions of healthy controls (mean age 40.8 y, range 21–70 y). As in the case of the static acquisitions, three cases had to be excluded due to metal artifacts caused by dental implants.

2.2 Imaging protocol

All patients were examined on a GE SIGNA TOF PET/MR (GE Healthcare, Chicago, US) [23]. The static [^{18}F]FDG PET/MR protocol included a 25 min static PET acquisition 66 ± 9 min after tracer injection (mean injected dose 144 ± 31 MBq). For the eleven [^{18}F]PE2I cases, 60 min of dynamic PET data were acquired directly after tracer injection (mean injected dose 153 ± 15 MBq). During the PET acquisitions a LAVA flex MR (acquisition details: repetition time 4 ms, echo time 2.23 ms, flip angle 5° , matrix $256 \times 256 \times 120$, voxel size $1.95 \text{ mm} \times 1.95 \text{ mm} \times 2.6 \text{ mm}$, number of averages 0.7, acquisition time: 18 s), a ZTE MR (acquisition details: 3D radial acquisition, flip angle 0.8° , matrix $110 \times 110 \times 116$, voxel size $2.4 \text{ mm} \times 2.4 \text{ mm} \times 2.4 \text{ mm}$, number of averages 4, bandwidth $\pm 62.5 \text{ kHz}$, acquisition time 42 s) and other study-specific MR sequences were acquired. Among the study-specific MR sequences were a 3D volumetric sagittal T1-weighted BRAVO sequence (acquisition details: echo time 3.2 ms, repetition time 8.5 ms, inversion time 450 ms, flip angle 12° , receiver bandwidth $\pm 31.2 \text{ kHz}$, NEX 1, voxel size $1 \text{ mm} \times 1 \text{ mm} \times 1 \text{ mm}$) and a 3D sagittal T2-weighted CUBE FLAIR sequence (acquisition details: echo time 137 ms, number of echoes 1, repetition time 8500 ms, inversion time 50 ms, receiver bandwidth $\pm 31.25 \text{ kHz}$, NEX 1, voxel size $1.2 \text{ mm} \times 1.3 \text{ mm} \times 1.4 \text{ mm}$). In all cases, a standard head coil (8-channel HR brain, GE Healthcare) was used for the MR acquisitions.

All subjects underwent a PET/CT acquisition before ([^{18}F]FDG cases) or after ([^{18}F]PE2I cases) the PET/MR acquisition. The PET/CT acquisitions were performed on a Siemens Biograph 16 or with a Siemens Biograph 40 (Siemens Healthcare, Erlangen, Germany) PET/CT. All PET/CT examinations included a low-dose CT acquisition (120 kV, 11 mAs) which was used to generate a CT-based attenuation image taken as the ground truth in the study.

2.3 PET image reconstruction

The PET raw data from all PET/MR acquisitions were reconstructed with three different methods for attenuation correction, shown in Figure 1. First, a GE atlas-based attenuation image (current default method in the SIGNA PET/MR) was used to reconstruct $\text{PET}_{\text{AtlasAC}}$. Subsequently, a GE ZTE-based attenuation image and a coregistered CT-based attenuation image were used to reconstruct $\text{PET}_{\text{ZTEAC}}$ and PET_{CTAC} , respectively. The generation of all attenuation images is described in detail in the following subsection. The reconstructions of the static PET data sets were performed offline with the GE reconstruction toolbox v.1.28 (GE Healthcare, Chicago, US).

using time of flight ordered subset maximum likelihood expectation maximization (TOF OSEM) with 4 iterations and 28 subsets, a voxel size of 1.17 mm x 1.17 mm x 2.78 mm, and a Gaussian post-smoothing with an FWHM of 4 mm.

Reconstruction of the dynamic [^{18}F]PE2I PET data sets was performed on the scanner (software version MP24.R03). The acquired listmode data were split into 32 frames (frame length 10 s to 360 s). All frames were reconstructed with TOF OSEM with 4 iterations and 28 subsets, a voxel size of 1.56 mm x 1.56 mm x 2.78 mm and a Gaussian post-smoothing with an FWHM of 3 mm.

2.4 Generation of attenuation images

First, the atlas-based attenuation images were generated with the GE reconstruction toolbox v.1.28 which uses the same post-processing algorithm as implemented in the current software release of the SIGNA PET/MR (MP24.R03). The algorithm uses a non-rigid registration of an input in-phase LAVA flex MR image to an atlas of predefined attenuation images [24, 25]. The resulting atlas-based attenuation images are post-smoothed with a Gaussian kernel with FWHM ca. 10 mm.

Second, the ZTE-based attenuation images were generated by post-processing the ZTE MR images with a research tool provided by GE (v.1.6.2). The upcoming software release of the SIGNA PET/MR (MP 26) will contain an option to use this algorithm for ZTE-based AC. The ZTE post-processing algorithm identifies bone voxels based on the ZTE image intensity and assigns continuous bone attenuation values [15, 17, 16]. The bone segmentation in the ZTE post-processing is completely model-free. To avoid misclassifications of air, tissue and bone in the nasal region, the ZTE post-processing algorithm v.1.6.2 uses the sinus/edge correction evaluated in [20].

The ZTE post-processing provided by GE has several input parameters. For all parameters but one (the partial volume slope) we used the default values suggested by GE. We used a value of 2 for the parameter for the partial volume slope which was obtained based on an evaluation of the results of the first 15 static subjects. The main influence of the partial volume slope parameter that we observed was a change in the size of the outer contour of the head (transition between background air and soft tissue of the skin). By changing the partial volume slope we obtained better agreement with the size of the outer contours derived from the CT-based attenuation images. When using the default partial volume slope of 1, the outer contour of the head is dilated by 1 voxel (2.4 mm) compared to using a partial volume slope of 2. This in turn led to a small global positive bias of 3%. All ZTE-based attenuation images were post-smoothed with a Gaussian kernel with FWHM 4 mm.

Third, after automatically removing the patient bed and cushions the low-dose CT images from the PET/CT acquisition were rigidly coregistered to the in-phase LAVA flex MR. Subsequently, the Hounsfield units of the coregistered CT were scaled to 511-keV attenuation coefficients by using the GE-provided multi-linear scaling. We verified that the Siemens and GE scaling for 120 kV are virtually identical up to 1200 HU (where GE decreases the slope while Siemens does not). After adding the templates for the PET/MR patient table and the head coil, a CT-based attenuation image that could be used to reconstruct the PET/MR raw data was obtained. All CT-based attenuation images were also post-smoothed with a Gaussian kernel with FWHM 4 mm.

The axial field of view (FOV) of the ZTE MR (limited by the sensitivity of the head coil) and the one of the CT was slightly smaller than the axial FOV of the PET detector rings in the SIGNA PET/MR. To complete areas in the neck and shoulders where ZTE or CT image information was




Figure 1. Workflow of the PET reconstructions used in this study. In the reconstruction of $PET_{AtlasAC}$ an atlas-based attenuation image that was derived from a LAVA flex MR image was used for attenuation correction (left column). The atlas-based attenuation image was generated with the vendor-provided software that is used in clinical routine. In the reconstruction of PET_{ZTEAC} attenuation correction was performed using a ZTE-based attenuation image that was derived from a ZTE MR image (middle column). The ZTE MR post-processing was done with a research tool provided by the vendor. For PET_{CTAC} a coregistered CT-based attenuation image from a PET/CT acquisition of the same day was used. In all attenuation images, templates for the bed and the head coil were added.

not available, a simple segmentation-based two class attenuation image based on the LAVA flex MR image was used.

2.5 Image analysis of static acquisitions

For all static acquisitions the mean uptake in 85 anatomical volumes of interest (VOIs) was calculated in $PET_{AtlasAC}$, PET_{ZTEAC} , and PET_{CTAC} . The VOIs were defined in the neuro tool of PMOD v.3.8 (PMOD technologies LCC, Zurich, Switzerland) using the Hammers atlas [26]. In every VOI we calculated the fractional bias of the mean uptake as

$$b_{AtlasAC}(VOI) = \frac{a_{AtlasAC}(VOI) - a_{CTAC}(VOI)}{a_{CTAC}(VOI)} \quad (1)$$

$$b_{ZTEAC}(VOI) = \frac{a_{ZTEAC}(VOI) - a_{CTAC}(VOI)}{a_{CTAC}(VOI)}, \quad (2)$$

where $a_{CTAC}(VOI)$ is the mean uptake of the VOI in PET_{CTAC} that was used as the gold standard and $a_{AtlasAC}(VOI)$ and $a_{ZTEAC}(VOI)$ are the mean uptake of the VOI in $PET_{AtlasAC}$ and PET_{ZTEAC} ,

respectively. In three subjects (6,7, and 21), the caudal end of the occipital skull was not completely in the FOV in the attenuation CT. In those subjects, the cerebellum VOIs were excluded from the analysis. All VOIs were grouped according to their anatomical location into the following groups: frontal cortex, temporal cortex, parietal cortex, occipital cortex, medial temporal cortex, striatum, thalamus, cerebellum, and cerebral white matter. All VOIs and the assigned groups (regions) are listed in supplementary Tables S1 and S2. A Wilcoxon signed-rank test was used to test whether the subject averaged mean of a_{AtlasAC} and a_{ZTEAC} is different from a_{CTAC} in all VOIs and regions.

To analyze the robustness of GE's atlas-based and ZTE-based attenuation correction, we applied the metric proposed in the multi-center evaluation of Ladefoged et al. [13]. This metric calculates the fraction of subjects in which the MRAC-introduced voxel bias of at least a given fraction of brain voxels is within $\pm 5\%$, $\pm 10\%$, $\pm 15\%$. As mentioned in [13], for a perfect AC method 100% of the subjects 100% of the brain voxels would be within $\pm 0\%$. The results of this metric were visualized in a characteristic curve for the three bias thresholds $\pm 5\%$, $\pm 10\%$, $\pm 15\%$. As in [13], we also analyzed three subjects with the biggest fraction of voxels exceeding a bias of $\pm 10\%$.

2.6 Image analysis of dynamic acquisitions

Regional time activity curves (TACs) were extracted for the left and right caudate nucleus, left and right putamen, and the cortex of the cerebellum. All VOIs were defined based on the 3D T1 BRAVO MR image using the Freesurfer image analysis suite which is documented and freely available online (<http://surfer.nmr.mgh.harvard.edu/>) [27]. Subsequently, we used the simplified reference tissue model (SRTM) with the cerebellar grey matter as reference region to estimate binding potential values (BP_{nd}) in the four striatal VOIs.

As proposed in [28] and validated for $[^{18}\text{F}]\text{PE2I}$ in [29], the tissue response $C_t(t)$ was modeled as

$$C_t(t) = R_1 C_r(t) + \left(k_2 - \frac{R_1 k_2}{1 + \text{BP}_{\text{nd}}} \right) C_r(t) * \exp \left(\frac{-k_2 t}{1 + \text{BP}_{\text{nd}}} \right), \quad (3)$$

where $C_r(t)$ is the TAC of the reference tissue (the cerebellum), R_1 is the ratio between K_1 of the tissue and reference tissue, and $*$ denotes the convolution operator. The parameters R_1 , k_2 , and BP_{nd} were obtained with non-linear curve fitting using the python package lmfit (v.0.9.7).

In a similar way to Eqs. (1) and (2), we calculated the bias of BP_{nd} , R_1 , and k_2 in the four striatal VOIs for $\text{PET}_{\text{AtlasAC}}$ and $\text{PET}_{\text{ZTEAC}}$ compared to PET_{CTAC} . In all VOIs, a Wilcoxon signed-rank test was used to test whether the subject averaged mean of $\text{BP}_{\text{nd,AtlasAC}}$ and $\text{BP}_{\text{nd,ZTEAC}}$ differ from $\text{BP}_{\text{nd,CTAC}}$.

3 RESULTS

3.1 Regional bias in static PET imaging

Figures 2 and 3, and supplementary Tables S3, and S4 show the results for the regional bias in the static $[^{18}\text{F}]\text{FDG}$ reconstructions caused by ZTEAC and AtlasAC compared to CTAC on a subject and regional level, respectively. Globally the bias ranges from ranges from -31.6% to +16.6% with a mean of -0.4% and a standard deviation of 4.3% for $\text{PET}_{\text{AtlasAC}}$. For $\text{PET}_{\text{ZTEAC}}$ the bias ranges from -8.0% to +7.7% with a mean of 0.1% and a standard deviation of 2.0%. Excluding the outliers

Table 1. Results of outlier analysis of the static acquisitions in terms of subjects with highest voxel bias, highest VOI bias and highest fraction of brain exceeding a bias of $\pm 10\%$

subjects with highest single voxel bias			
AtlasAC	subject 24 -53% (left straight gyrus)	subject 29 -43% (left anterior temporal lobe)	subject 21 +41% (left superior frontal gyrus)
ZTEAC	subject 5 +45% (left cerebellum)	subject 18 +43% (left cerebellum)	subject 21 +41% (right fusiform gyrus)
subjects with highest VOI bias			
AtlasAC	subject 24 -31% (left straight gyrus)	subject 22 +16% (left superior frontal gyrus)	subject 21 +16% (left precentral gyrus)
ZTEAC	subject 9 -8% (left middle frontal gyrus)	subject 22 +7% (right superior temporal gyrus)	subject 21 +7% (right fusiform gyrus)
subjects with highest fraction of brain exceeding a bias of $\pm 10\%$			
AtlasAC	subject 22 23%	subject 19 22%	subject 21 18%
ZTEAC	subject 21 5%	subject 6 5%	subject 9 3%

193 based on the boxplot shown in Fig. 3 reduces the global bias range to -12% to $+14\%$ for PET_{AtlasAC}
 194 and to -5.5% to $+5.5\%$ for PET_{ZTEAC}.

195 On a subject level, Fig. 2 and supplementary Table S3 demonstrate that ZTEAC strongly reduces
 196 the inter- and intra-subject variability in the bias. In subject 24 where the non-rigid alignment to
 197 the atlas failed, PET_{AtlasAC} showed severe negative bias of more than -25% in the orbitofrontal
 198 cortical VOIs (see Fig. 5). In these VOIs of subject 24, the bias of PET_{ZTEAC} was less than 1.6% .

199 On a regional level, Fig. 3 and supplementary Table S4 show that ZTEAC strongly reduces the
 200 inter- and intra-regional variability in the bias, as well as the mean bias in the frontal cortex,
 201 temporal cortex, parietal cortex, medial temporal cortex, cerebellum, and cerebral white matter. In
 202 all regions shown in Fig. 2, the mean bias in PET_{ZTEAC} is between -1.2% and $+0.6\%$. PET_{AtlasAC}
 203 shows a distinct negative bias in the cerebellum (mean -4.2%) and distinct positive bias in the
 204 parietal cortex (mean $+4\%$).

205 On a VOI level, Fig. 4 and supplementary Tables S1 and S2 demonstrate that ZTEAC strongly
 206 reduces the inter- and intra-VOI variability in the bias, as well as the mean bias in almost all
 207 VOIs. The mean VOI bias caused by ZTEAC ranges from -1.8% in the lateral remainder of the left
 208 occipital lobe to $+2.2\%$ in the left lateral ventricle. In PET_{AtlasAC}, a distinct gradient in the mean
 209 VOI bias in the cranio-caudal direction is visible. The mean VOI bias caused by AtlasAC ranges
 210 from -4.5% in the cerebral white matter to $+6.6\%$ in the left superior frontal gyrus. In PET_{ZTEAC},
 211 only 1.4% of the analyzed VOIs in all subjects had a bias of more than 5% whereas in PET_{AtlasAC}
 212 20.3% of all VOIs showed a bias of more than 5% .

213 Figure 6 shows the results of the outlier metric [13] for biases within ($\pm 5\%$, $\pm 10\%$, $\pm 15\%$). Again,
 214 the performance of ZTE-based attenuation correction is much better than the one of the atlas-based
 215 attenuation correction. At least 95% / 77% of all brain voxels in all subjects show a bias within
 216 $\pm 10\%$ for PET_{ZTEAC} / PET_{AtlasAC}. For a bias within $\pm 5\%$ the corresponding values are 82% /
 217 46% and for a bias within $\pm 15\%$ the corresponding values are 97% / 89% . Table 1 shows results
 218 of three worst outliers in terms of subjects with highest voxel bias, highest VOI bias and highest
 219 fraction of brain exceeding a bias of $\pm 10\%$.




Figure 2. Regional bias in the PET reconstruction caused by AtlasAC (blue) and ZTEAC (red) compared to CTAC for all 30 static PET acquisitions. Each box plot shows the bias distribution over the 85 anatomical VOIs. The rectangular boxes represent the interquartile ranges (IQR) and the horizontal line are the medians. The upper ends of the whiskers are at the minimum of: the third quartile plus 1.5IQR and the biggest data point. The lower ends of the whiskers are at the maximum of: the first quartile minus 1.5IQR and the smallest data point. Outliers are plotted with open circles. Please note that the VOIs in the cerebellum had to be excluded in 3 subjects (6,7,21).




Figure 3. Regional bias in the PET reconstruction caused by AtlasAC (blue) and ZTEAC (red) compared to CTAC as a function of the VOI location in the brain. Please note that the VOIs in the cerebellum had to be excluded in 3 subjects (6,7,21).




Figure 4. (a) mean of bias in the PET reconstruction in all 85 anatomical VOIs averaged over all 30 static PET acquisitions for Atlas-AC (top row) and ZTE-AC (bottom row). (b) standard deviation of VOI bias. The VOI location is visualized in five sagittal slices using the brain anatomy of subject 1. Please note that the VOIs in the cerebellum had to be excluded in 3 subjects (6,7,21).

3.2 Bias in kinetic modeling of $[^{18}\text{F}]\text{PE2I}$

Figure 7 and supplementary Table S5 summarize the bias in the modeled BP_{nd} in four different regions of the striatum using the cerebellum as reference region and TACs derived from $\text{PET}_{\text{AtlasAC}}$ and $\text{PET}_{\text{ZTEAC}}$ compared to TACs from PET_{CTAC} . The bias in the BP_{nd} ranges from -0.6% (right putamen in subject 11) to +9.4% (left caudate nucleus subject 9) for $\text{PET}_{\text{AtlasAC}}$ and from -0.8% (right putamen subject 3) to +4.8% (right caudate nucleus subject 8) for $\text{PET}_{\text{ZTEAC}}$. The right caudate nucleus shows the biggest subject averaged regional bias ($5.1\% \pm 2.6\%$, $p = 0.003$ for $\text{PET}_{\text{AtlasAC}}$ and $2.0\% \pm 1.5\%$, $p = 0.006$ for $\text{PET}_{\text{ZTEAC}}$). In addition, supplementary Figs. S1, S2, and S3 show the bias in the time activity curves, and the R_1 and k_2 estimates in $\text{PET}_{\text{AtlasAC}}$ and $\text{PET}_{\text{ZTEAC}}$, respectively.

4 DISCUSSION

Our analysis demonstrates that the bias caused by ZTEAC compared to CTAC as ground truth for brain PET/MR is small. The magnitude of the maximum bias of 8% is in agreement with the




Figure 5. Transversal slices of (a) atlas-based attenuation image, (b) ZTE-based attenuation image, (c) CT-based attenuation image, (d) regional bias in PET_{AtlasAC}, and (e) regional bias in PET_{ZTEAC} of subject 24. In this case, the template registration in the atlas-based attenuation image failed which caused a misclassification of soft tissue voxels as air voxels in the frontal region. The cyan contour lines show the head contour in the CT-based attenuation image for comparison. As a result of the underestimated attenuation PET_{AtlasAC} shows strong negative bias of up to -32% in the left straight gyrus.

analysis of Sekine et al. [16]. In contrast to [16], we have evaluated more static PET as well as dynamic PET acquisitions. Moreover, the subjects in our analysis underwent a PET/MR protocol with realistic PET acquisition times whereas [16] only used an additional two minute PET/MR exam after a PET/CT acquisition.

In contrast to the earlier evaluation of the quantitative accuracy of the AtlasAC [24, 25] (range of VOI bias -5% to +7.3%), our analysis showed that the AtlasAC implemented in the GE SIGNA PET/MR can lead to individual regional underestimations of up to -32% (as observed in subject 24). A possible reason for the discrepancy is the fact that the number of subjects in [24, 25] was much smaller compared to our study. In this work, the biggest underestimations were found in a single subject (24) where the alignment of the atlas to the patient anatomy failed (see Fig. 5) which caused a misclassification of some soft-tissue voxels as air voxels (pharynx) in the frontal region. Since the atlas alignment is highly subject dependent, failures are hard to predict. As demonstrated in subject 24, those failures can occur with the current implementation of the AtlasAC leading to severe problems in regional quantification.

As observed in [16], another drawback of the current AtlasAC is the fact that the introduced bias in the PET reconstruction shows a clear gradient in the cranio-caudal direction. Caudal VOIs such as the cerebellum (ca -4.2%) and the anterior lateral temporal lobe (-4.2%) show negative bias. This is because part of the temporal and occipital bone are classified as soft tissue in the current AtlasAC. Moreover, there is a gross underestimation of the anterior part of the head including the oropharynx, nasal cavities and cartilage tissue.

On the other hand, the superior cortical areas (frontal-parietal) show positive bias (+6.6% in the left superior frontal gyrus). This overestimation is caused by the fact that (a) the thickness of




Figure 6. Outlier analysis [13] of the 30 static acquisitions for PET_{ZTEAC} (red) and PET_{AtlasAC} (blue). Note the different scale on the x-axis.

the superior skull seems overestimated in the AtlasAC and (b) the atlas-based attenuation image is heavily post-smoothed such that some soft tissue voxels in the superior gyri in the attenuation image are affected by spill over from skull voxels.

The cranio-caudal gradient in the bias distribution affects especially cerebral kinetic modeling analysis when using the cerebellum as the reference region. This could be demonstrated in the kinetic modeling of the binding potential in the striatum of the eleven [¹⁸F]PE2I subjects. As a consequence of the observed negative bias in the cerebellum compared to the striatum in PET_{AtlasAC}




Figure 7. (A) Boxplot of BP_{nd} values in four striatal regions of the 11 [¹⁸F]PE2I subjects obtained from PET_{AtlasAC}, PET_{ZTEAC}, and PET_{CTAC}. (B) Bias in BP_{nd} estimation in PET_{AtlasAC}, PET_{ZTEAC} compared to PET_{CTAC}.

in the static cases, AtlasAC leads to a small but systematic and significant overestimation of the binding potential of [¹⁸F]PE2I in the striatum (ca. +5% in the caudate nucleus and +3.3% in the putamen). This positive bias can be understood by looking at Eq. (3). Under the assumptions that BP_{nd} ≫ 1 and 1 + BP_{nd} ≫ R₁ (which both are fulfilled for [¹⁸F]PE2I in the striatum), it can be seen that scaling C_r(t) with α and at the same time scaling R₁, k₂, and BP_{nd} with α⁻¹ yields the exact same tissue response C_t(t). Since we can deduce from the analysis of the static examinations that C_r(t) in the cerebellum is underestimated by ca. 4% and that there is almost no bias in the striatum (C_t(t)), we would expect a 4% overestimation in R₁, k₂, and BP_{nd} which is in accordance with the results of the dynamic analysis as shown in Fig. 7, and supplementary Figs S2 and S3.

Using ZTEAC strongly reduces this bias in BP_{nd} (ca. +2.0% in the caudate nucleus and +1.1% in the putamen). The performance of ZTEAC in the context of dynamic PET imaging is comparable to the MaxProb multi atlas-based attenuation correction method. In [30], Merida et al. could show that the MaxProb method leads to a regional bias of -2% to +5% in the BP_{nd} of seven subjects examined with [¹⁸F]MPPF.

It has been shown [25, 31] and should be noted that the AtlasAC method implemented in the GE SIGNA PET/MR is clearly outperformed by more advanced atlas-based methods. Since the focus of this study was to analyze the performance of the ZTE-based attenuation correction that will become clinically available on the SIGNA PET/MR, a detailed analysis of more advanced atlas-based methods for MR-based attenuation correction is beyond the scope of this study.

Compared to the detailed multi-center study of 11 methods for brain attenuation correction for the Siemens mMR in 359 subjects by Ladefoged et al. [13], it can be seen that the results for the regional quantitative accuracy of ZTEAC are comparable to the best methods in [13] which showed a global mean bias in the range of -0.4% to +0.8% with a standard deviation of 1.2% to 1.9%. Also in terms of robustness (as seen in the standard deviation in the VOI-averaged bias) and in terms of outlier behavior ZTEAC performs comparably to to best methods of [13]. However, it should be noted that we could only analyze 30 subjects which influences the detection of (rare) outliers.

Among the five best methods in [13] are three template-/atlas-based methods [8, 31, 12] and two ultra short echo time MR (UTE) segmentation-based methods [4, 5]. Compared to the template-/atlas-based methods, the current ZTEAC for brain has the advantage that it does not rely on any anatomical prior information. This might be beneficial in subjects with very abnormal brain anatomy (e.g. after surgery or traumatic brain injury) which needs further validation.

Finally, the fact that we had to exclude 6 out of 48 patients (12.5%) due MR artifacts caused by dental implants demonstrates that there is a need for a reliable method for compensation of metal artifacts that can be applied in clinical routine.

A potential limitation of the study is the fact that the attenuation CTs used for CTAC were acquired on a Siemens PET/CT system, but scaled to linear attenuation coefficients with the multi-linear scaling provided by GE. This might lead to small residual uncertainties in the linear attenuation coefficients of the ground truth CTAC due to the fact that the vendor-specific scaling procedures might be optimized for different effective x-ray spectra. However, we do not expect this to be a major problem, because the multi-linear scaling curves of GE and Siemens are virtually identical up to 1200 HU.

5 CONCLUSION

ZTE-based attenuation correction provides excellent quantitative accuracy for static and dynamic PET/MR imaging in all parts of the brain. It is clearly superior to the Atlas-based head attenuation correction currently implemented in the GE SIGNA PET/MR and hereby obviates the major concern that was present in the quantitative accuracy of brain PET/MR.

6 DECLARATIONS

List of abbreviations

PET	positron emission tomography
MR	magnetic resonance imaging
AC	attenuation correction
MRAC	MR-based AC
ZTE MR	zero echo time MR
ZTEAC	ZTE MR-based AC
CTAC	CT-based AC
AtlasAC	Atlas-based AC
UTE MR	ultra short echo time MR
HU	Hounsfield unit
TOF	time of flight
OSEM	ordered subset maximum likelihood expectation maximization
FOV	field of view
VOI	volume of interest
TAC	time activity curve

Ethics approval and consent to participate

All procedures performed in studies involving human participants were in accordance with the ethical standards of the institutional and/or national research committee and with the 1964 Helsinki declaration and its later amendments or comparable ethical standards. Informed consent was obtained from all individual participants included in the study.

Consent for publication

Informed consent was obtained from all individual participants included in the study.

Availability of data and material

Please contact author for data requests.

Funding

GS is funded by the NIH project 1P41EB017183-01A1. AR and JN are funded by the Research Foundation Flanders (FWO) projects 12T7118N and G.0275.14N.

Authors' contributions

GS performed the PET reconstructions, the static PET image analysis and drafted the manuscript. MK performed the analysis of the dynamic PET images. SW coordinated the study and helped to draft the manuscript. AR helped to perform the PET reconstructions. DVW coordinated the study and helped to draft the manuscript. GD performed the postprocessing of the ZTE MR images. RP and NM optimized the ZTE MR sequence on the SIGNA PET/MR at UZ Leuven. JN and KVL designed the study and helped to draft the manuscript.

Acknowledgements

Not applicable.

COMPETING INTERESTS

GD is an employee of GE Healthcare, Cambridge, UK. GS, MK, AR, NM, RP, JN, and KVL received travel grants from GE Healthcare for attending workshops. GS and KVL received travel grants for presenting PET/MR results at GE user meetings. KU Leuven received a research collaboration grant for partial funding of this study. All other authors declare that they have no conflict of interest.

REFERENCES

- [1] Andersen FL, Ladefoged CN, Beyer T, Keller SH, Hansen AE, Højgaard L, et al. Combined PET/MR imaging in neurology: MR-based attenuation correction implies a strong spatial bias when ignoring bone. *Neuroimage* **84** (2014) 206–16. doi:10.1016/j.neuroimage.2013.08.042.
- [2] Keereman V, Fierens Y, Broux T, De Deene Y, Lonnew M, Vandenberghe S. MRI-based attenuation correction for PET/MRI using ultrashort echo time sequences. *J Nucl Med* **51** (2010) 812–8. doi:10.2967/jnumed.109.065425.

- 340 [3] Delso G, Carl M, Wiesinger F, Sacolick L, Porto M, Hullner M, et al. Anatomic Evaluation of
341 3-Dimensional Ultrashort-Echo-Time Bone Maps for PET/MR Attenuation Correction. *J Nucl*
342 *Med* **55** (2014) 780–785. doi:10.2967/jnumed.113.130880.
- 343 [4] Ladefoged CN, Benoit D, Law I, Holm S, Kjær A, Højgaard L, et al. Region specific optimization
344 of continuous linear attenuation coefficients based on UTE (RESOLUTE): application to
345 PET/MR brain imaging. *Phys Med Biol* **60** (2015) 8047–8065. doi:10.1088/0031-9155/60/20/
346 8047.
- 347 [5] Juttukonda MR, Mersereau BG, Chen Y, Su Y, Rubin BG, Benzinger TL, et al. MR-based
348 attenuation correction for PET/MRI neurological studies with continuous-valued attenuation
349 coefficients for bone through a conversion from R2* to CT-Hounsfield units. *Neuroimage* **112**
350 (2015) 160–168. doi:10.1016/j.neuroimage.2015.03.009.
- 351 [6] Anazodo UC, Thiessen JD, Ssali T, Mandel J, Günther M, Butler J, et al. Feasibility of
352 simultaneous whole-brain imaging on an integrated PET-MRI system using an enhanced 2-point
353 Dixon attenuation correction method. *Front Neurosci* **9** (2015) 1–11. doi:10.3389/fnins.2014.
354 00434.
- 355 [7] Cabello J, Lukas M, Rota Kops E, Ribeiro A, Shah NJ, Yakushev I, et al. Comparison between
356 MRI-based attenuation correction methods for brain PET in dementia patients. *Eur J Nucl*
357 *Med Mol Imaging* **43** (2016) 2190–2200. doi:10.1007/s00259-016-3394-5.
- 358 [8] Izquierdo-Garcia D, Hansen AE, Forster S, Benoit D, Schachoff S, Furst S, et al. An SPM8-
359 Based Approach for Attenuation Correction Combining Segmentation and Nonrigid Template
360 Formation: Application to Simultaneous PET/MR Brain Imaging. *J Nucl Med* **55** (2014)
361 1825–1830. doi:10.2967/jnumed.113.136341.
- 362 [9] Paulus DH, Quick HH, Geppert C, Fenchel M, Hermosillo G, Faul D, et al. Whole-Body PET /
363 MR Imaging : Quantitative Evaluation of a Novel Model-Based MR Attenuation Correction
364 Method Including Bone. *J Nucl Med* **56** (2015) 1061–1066. doi:10.2967/jnumed.115.156000.
- 365 [10] Koesters T, Friedman KP, Fenchel M, Zhan Y, Hermosillo G, Babb J, et al. Dixon Sequence
366 with Superimposed Model-Based Bone Compartment Provides Highly Accurate PET/MR
367 Attenuation Correction of the Brain. *J Nucl Med* **57** (2016) 918–924. doi:10.2967/jnumed.115.
368 166967.
- 369 [11] Burgos N, Cardoso MJ, Thielemans K, Modat M, Pedemonte S, Dickson J, et al. Attenuation
370 correction synthesis for hybrid PET-MR scanners: Application to brain studies. *IEEE Trans*
371 *Med Imaging* **33** (2014) 2332–2341. doi:10.1109/TMI.2014.2340135.
- 372 [12] Merida I, Costes N, Heckemann R, Hammers A. Pseudo-CT generation in brain MR-PET
373 attenuation correction: comparison of several multi-atlas methods. *IEEE 12th Int Symp Biomed*
374 *Imaging* (2015), vol. 1, 1431–1434. doi:10.1186/2197-7364-2-S1-A29.
- 375 [13] Ladefoged CN, Law I, Anazodo U, St Lawrence K, Izquierdo-Garcia D, Catana C, et al. A
376 multi-centre evaluation of eleven clinically feasible brain PET/MRI attenuation correction
377 techniques using a large cohort of patients. *Neuroimage* **147** (2017) 346–359. doi:10.1016/j.
378 neuroimage.2016.12.010.
- 379 [14] Weiger M, Stampanoni M, Pruessmann KP. Direct depiction of bone microstructure using MRI
380 with zero echo time. *Bone* **54** (2013) 44–47. doi:10.1016/j.bone.2013.01.027.
- 381 [15] Wiesinger F, Sacolick LI, Menini A, Kaushik SS, Ahn S, Veit-Haibach P, et al. Zero TE MR
382 bone imaging in the head. *Magn Reson Med* **114** (2015) 107–114. doi:10.1002/mrm.25545.

- [16] Sekine T, ter Voert EEGW, Warnock G, Buck A, Huellner M, Veit-Haibach P, et al. Clinical Evaluation of Zero-Echo-Time Attenuation Correction for Brain 18F-FDG PET/MRI: Comparison with Atlas Attenuation Correction. *J Nucl Med* **57** (2016) 1927–1932. doi:10.2967/jnumed.116.175398.
- [17] Delso G, Wiesinger F, Sacolick LI, Kaushik SS, Shanbhag DD, Hullner M, et al. Clinical Evaluation of Zero-Echo-Time MR Imaging for the Segmentation of the Skull. *J Nucl Med* **56** (2015) 417–422. doi:10.2967/jnumed.114.149997.
- [18] Boydev C, Demol B, Pasquier D, Saint-Jalmes H, Delpon G, Reynaert N. Zero echo time MRI-only treatment planning for radiation therapy of brain tumors after resection. *Phys Medica* **42** (2017) 332–338. doi:10.1016/j.ejmp.2017.04.028.
- [19] Khalifé M, Fernandez B, Jaubert O, Soussan M, Brulon V, Buvat I, et al. Subject-specific bone attenuation correction for brain PET/MR: can ZTE-MRI substitute CT scan accurately? *Phys Med Biol* **62** (2017) 7814–7832. doi:10.1088/1361-6560/aa8851.
- [20] Yang J, Wiesinger F, Kaushik S, Shanbhag D, Hope TA, Larson PEZ, et al. Evaluation of sinus/edge corrected ZTE-based attenuation correction in brain PET/MRI. *J Nucl Med* (2017). doi:10.2967/jnumed.116.188268.
- [21] Leynes AP, Yang J, Hope TA. Hybrid ZTE / Dixon MR-based attenuation correction for quantitative uptake estimation of pelvic lesions in PET / MRI. *Med Phys* **44** (2017) 902–913.
- [22] Wiesinger F, Bylund M, Yang J, Kaushik S, Shanbhag D, Ahn S, et al. Zero TE-based pseudo-CT image conversion in the head and its application in PET/MR attenuation correction and MR-guided radiation therapy planning. *Magn Reson Med* **00** (2018) 1–12. doi:10.1002/mrm.27134.
- [23] Grant AM, Deller TW, Khalighi MM, Maramraju SH, Delso G, Levin CS. NEMA NU 2-2012 performance studies for the SiPM-based ToF-PET component of the GE SIGNA PET/MR system. *Med Phys* **43** (2016) 2334–2343. doi:10.1118/1.4945416.
- [24] Sekine T, Buck A, Delso G, ter Voert EEGW, Huellner M, Veit-Haibach P, et al. Evaluation of Atlas-Based Attenuation Correction for Integrated PET/MR in Human Brain: Application of a Head Atlas and Comparison to True CT-Based Attenuation Correction. *J Nucl Med* **57** (2016) 215–220. doi:10.2967/jnumed.115.159228.
- [25] Sekine T, Burgos N, Warnock G, Huellner M, Buck A, ter Voert EEGW, et al. Multi-Atlas-Based Attenuation Correction for Brain 18F-FDG PET Imaging Using a Time-of-Flight PET/MR Scanner: Comparison with Clinical Single-Atlas- and CT-Based Attenuation Correction. *J Nucl Med* **57** (2016) 1258–1264. doi:10.2967/jnumed.115.169045.
- [26] Hammers A, Allom R, Koeppe MJ, Free SL, Myers R, Lemieux L, et al. Three-dimensional maximum probability atlas of the human brain, with particular reference to the temporal lobe. *Hum Brain Mapp* **19** (2003) 224–247. doi:10.1002/hbm.10123.
- [27] Fischl B, Salat DH, Busa E, Albert M, Dieterich M, Haselgrove C, et al. Whole brain segmentation: automated labeling of neuroanatomical structures in the human brain. *Neuron* **33** (2002) 341–355.
- [28] Lammertsma AA, Hume SP. Simplified Reference Tissue Model for PET Receptor Studies. *Neuroimage* **4** (1996) 153–158. doi:10.1006/nimg.1996.0066.
- [29] Sasaki T, Ito H, Kimura Y, Arakawa R, Takano H, Seki C, et al. Quantification of Dopamine Transporter in Human Brain Using PET with 18F-FE-PE2I. *J Nucl Med* **53** (2012) 1065–1073. doi:10.2967/jnumed.111.101626.

- 427 [30] Merida I, Reilhac A, Redoute J, Heckemann R, Costes N, Hammers A. Multi-atlas attenuation
428 correction supports full quantification of static and dynamic brain PET data in PET-MR. *Phys*
429 *Med Biol* **62** (2017) 2834–2858.
- 430 [31] Burgos N, Cardoso MJ, Thielemans K, Modat M, Dickson J, Schott JM, et al. Multi-contrast
431 attenuation map synthesis for PET/MR scanners: assessment on FDG and Florbetapir PET
432 tracers. *Eur J Nucl Med Mol Imaging* **42** (2015) 1447–1458. doi:10.1007/s00259-015-3082-x.