

Administrative Data
Institution Samsung Medical Center
Physicist JKH
PatientID 56033983
Patient Name CHANG SANG LYOOL, 104659
Comment

Gamma 2D - Parameters
2.0 mm Distance- To- Agreement
2.0 % Dose difference with ref. to max. dose of calculated volume
Suppress dose below 5.0 % of max. dose of calculated volume
Option "Use 2nd and 3rd pass" selected

Statistics
Number of Dose Points 210
Evaluated Dose Points 87 (41.4 %)
Passed 87 (100.0 %)
Failed 0 (0.0 %)
Result 100.0 % ● (Green)

Gamma 2D
Arithmetic Mean 0.295
Min (LR = -5.0 mm; TG = -5.0 mm) 0.005
Max (LR = 0.0 mm; TG = 10.0 mm) 0.892
Median 0.265

Absolute Difference
Arithmetic Mean 0.026 Gy
Min (LR = -5.0 mm; TG = -5.0 mm) 0.000 Gy
Max (LR = 30.0 mm; TG = -10.0 mm) 0.075 Gy
Median 0.019 Gy

Settings
Passing criteria Gamma ≤ 1.0
Green 90.0 % to 100.0 %
Yellow 75.0 % to 90.0 %
Red 0.0 % to 75.0 %

