

Site Data:

Institution: Hospital Oncológico Dr. Miguel Pérez Carreño

Performed by: Lic. Nelson Sandoval Puente

Measurement Date: 2025-05-16

Date of Report: Wednesday, June 03, 2026

Unit: Elekta Precise

Energy: 6 MV

Instrumentation:

Chamber: 30013

Chamber Calibration Factor N_{dw} (cGy/nC): 5.443

Electrometer:

Kelec: 1.000

MU: 200

Beam Quality:

TPR(20)/TPR(10): 0.574

Determined kQ: 1.000

Chamber Corrections/Measurements:

Temperature (°C): 22.1

Pressure (kPa): 100.7

Mraw @ (-300V, Reference) (nC): (25.65, 25.66, 25.65)

Mraw @ (-150V, Reduced) (nC): (25.64, 25.63, 25.63)

Mraw @ (300V, Opposite) (nC): (25.64, 25.65, 25.65)

Ktp: 1.014

Ks: 1.001

Kpol: 1.000

Dose Determination:

Fully corrected M (nC): 26.026

Tissue correction (e.g. muscle): 1.000

Dose/MU @ z_{ref} depth (cGy): 0.708

Clinical PDD (%): 66.50

Dose/MU @ z_{max} (cGy): 1.065

Output Adjustment?: No

Notes:

note 1 from toml

note 2 from toml