



## World Expert Meeting in Arthroplasty 2024

## Is There a Difference in the Outcome of Primary Total Knee Arthroplasty When Mobile-Bearing Versus Fixed-Bearing Implants Are Used?



Quanjun Cui, MD <sup>a</sup>, Zhichang Zhang, MD <sup>b</sup>, Elizabeth Driskill, MS <sup>a</sup>,  
Corinne Vennitti, MD <sup>a</sup>, Charles Engh, MD <sup>a</sup>, John F. Burke, MD <sup>a</sup>, Abtin Alvand, MD <sup>c</sup>,  
Ahmad Abbaszadeh, MD <sup>d</sup>, Wendy Novicoff, PhD <sup>a,\*</sup>

<sup>a</sup> Department of Orthopaedic Surgery, University of Virginia, Charlottesville, Virginia

<sup>b</sup> Department of Orthopaedic Surgery, The First Affiliated Hospital of Xinxiang Medical University, Weihui, Henan, China

<sup>c</sup> Nuffield Department of Orthopaedics, Rheumatology & Musculoskeletal Sciences, University of Oxford, Oxford, UK

<sup>d</sup> Joint Reconstruction Research Center, Tehran University of Medical Sciences, Tehran, Iran

## ARTICLE INFO

## Article history:

Received 19 September 2024

Received in revised form

10 October 2024

Accepted 14 October 2024

Available online 28 October 2024

## Keywords:

total knee arthroplasty

mobile-bearing

fixed-bearing

outcomes

revision rate

aseptic loosening

**Question:** Is there a difference in the outcome of primary total knee arthroplasty when mobile-bearing versus fixed-bearing implants are used?

**Response/Recommendation:** The literature supports the notion that both mobile-bearing and fixed-bearing designs can achieve excellent outcomes. The theoretical advantages of mobile-bearing designs over fixed-bearing designs have not been substantiated in clinical practice. Consequently, we recommend that orthopaedic surgeons select mobile-bearing or fixed-bearing inserts based on their personal preferences and clinical judgment.

**Faculty Votes:** Agree: 89.9%, Disagree: 5.3%, Abstain: 4.8%.

**Level of Evidence:** moderate.

## Rationale

Fixed-bearing (FB) implants were introduced first and continue to represent the most common type used in total knee arthroplasty (TKA) [1,2]. These implants provide rigid fixation of the

One or more of the authors of this paper have disclosed potential or pertinent conflicts of interest, which may include receipt of payment, either direct or indirect, institutional support, or association with an entity in the biomedical field which may be perceived to have potential conflict of interest with this work. For full disclosure statements refer to <https://doi.org/10.1016/j.arth.2024.10.084>.

This research utilized the web for literature search and analysis. Our institutional review board deemed this study exempt.

This work was performed at the Department of Orthopaedic Surgery, University of Virginia School of Medicine, Charlottesville, VA, USA.

\* Address correspondence to: Wendy Novicoff, PhD, University of Virginia School of Medicine, 2280 Ivy Road, Charlottesville, VA 22903.

<https://doi.org/10.1016/j.arth.2024.10.084>

0883-5403/© 2024 Elsevier Inc. All rights are reserved, including those for text and data mining, AI training, and similar technologies.

polyethylene insert within the tibial implant and have demonstrated satisfactory outcomes and long-term survival rates [3,4]. However, implant loosening in FB designs has been theoretically attributed to higher contact stresses and polyethylene wear rates [5–7]. In response to these concerns, mobile-bearing (MB) polyethylene designs were developed. These designs aim to mitigate the drawbacks of FB TKA by improving conformity and lowering contact stresses, thereby mimicking the kinematics of the knee and promoting a wider range of motion [8,9]. However, these advantages remain theoretical and have yet to be fully proven *in vivo*. Furthermore, MB TKA can introduce unique complications, such as bearing dislocation [10].

Initial evidence suggested that MB designs might promote better outcomes in functional scores and complications, but the differences observed were minimal [11]. Subsequent meta-analyses with mid-term follow-up refuted these findings, indicating no significant difference between MB and FB TKA [12–14]. However,

recent studies have presented further contradictory results, and controversy continues regarding the superiority of MB over FB designs [15–18]. We conducted an updated review comparing MB versus FB TKA to address this ongoing debate. This analysis utilized a multimodal approach to outcomes, including overall revision rates, aseptic loosening, knee function scores, and radiological outcomes.

## Search Strategy

A comprehensive search in PubMed, Embase, Web of Science, and the Cochrane Library databases was done to identify randomized controlled trials (RCTs) that compared MB to FB prosthesis designs. The search was conducted according to the Preferred Reporting Items for Systematic Reviews and Meta-Analyses: the 2020 Preferred Reporting Items for Systematic Reviews and Meta-Analyses statement [19]. Search terms included “total knee arthroplasty” AND “mobile bearing” AND “fixed bearing” AND “randomized controlled trials.” The search was limited to RCT studies written in English that were published in the year 2000 or later and had at least two years of follow-up.

The search strategy resulted in 1,244 studies (1,240 from the database search and four articles from a manual reference search), of which 678 were excluded due to duplications. A total of 566 studies were screened by titles and abstracts, which resulted in 449 additional studies being excluded. A total of 117 studies were sought for retrieval and underwent full-text reviews, which led to an additional 47 studies being excluded. A total of 70 articles were included in this systematic review.

## Functional Outcomes

The comparative analysis of functional outcomes between FB and MB TKA reveals no significant long-term differences. Multiple studies have demonstrated no differences in patient-reported outcomes at two or three years postoperatively [20–39]. Similarly, evaluations at 10 years postoperatively show no significant differences in Knee Society Scores, knee flexion, or pain scores between the two types of prostheses [4,40–42]. Although several studies reported better outcomes for MB prostheses in the short term, this advantage was not sustained at three and five years, suggesting only temporary benefits [43–45]. Moreover, gait analysis and other functional metrics, such as range of motion and joint awareness, show no clear superiority of MB over FB TKAs in the mid to long term [3,8,16,44,46–66]. These findings indicate that the theoretical benefits of MB TKAs, such as improved kinematics and reduced wear, do not translate into significant clinical advantages for patients over FB TKAs in routine activities.

## Radiological Outcomes and Implant Migration

Radiological outcomes and implant migration assessments show no significant differences between FB and MB TKAs over various follow-up periods [27,31,53,67]. Several studies indicate similar migration patterns and rates of osteolysis for both types of implants [8,42,68–70]. Long-term follow-ups reveal no significant differences in the prevalence of radiolucent lines or osteolysis, further supporting the radiological equivalence of the two prosthesis types [4,71,72]. Although there were occasional reports of higher migration rates for FB tibial implants compared to MB implants, these differences were not clinically significant and did not affect the overall survivorship or functional outcomes of the prostheses [23,41,73,74]. These consistent findings across various studies suggest that the choice between FB and MB TKAs does not

impact radiological outcomes or the likelihood of implant loosening and migration.

## Implant Survival

Due to the limitation of sample volumes in RCTs, revisions were often reported as individual cases rather than as revision rates in some studies [48,51,54,66,75,76]. However, the implant survival data further support the equivalence of FB and MB TKAs. Studies with long-term follow-ups demonstrate no significant differences in overall survivorship between the two types of implants [42,71]. Although there were isolated reports of higher revision rates in MB TKAs due to aseptic loosening and other complications, these instances were relatively rare and did not significantly affect the overall survival rates of the prostheses [17,40,77]. High-flexion MB prostheses, while showing some early functional advantages, did not exhibit superior long-term survivorship compared to FB counterparts [3,4,22]. The comparable complication rates and implant survival outcomes across multiple studies indicate that both FB and MB TKAs are reliable options with similar long-term durability and safety profiles.

In conclusion, both MB and FB designs can achieve excellent outcomes. The choice between fixed and mobile bearing implants for TKA should be based on surgeon preference and experience with the selected implant, as the existing evidence does not indicate significant differences in functional outcomes, radiological results, complication rates, or implant survival between the two prosthesis types. Additional research to investigate long-term outcomes in real-world usage should be done using information from longitudinal registries.

## CRediT authorship contribution statement

**Quanjun Cui:** Writing – review & editing, Validation, Supervision, Project administration, Methodology, Conceptualization. **Zhichang Zhang:** Writing – original draft, Methodology, Investigation. **Elizabeth Driskill:** Writing – review & editing, Investigation, Formal analysis, Data curation. **Corinne Vennitti:** Writing – review & editing, Investigation, Formal analysis, Data curation. **Charles Engh:** Writing – review & editing, Investigation, Formal analysis, Data curation. **John F. Burke:** Writing – review & editing, Investigation, Formal analysis, Data curation. **Abtin Alvand:** Writing – review & editing, Validation, Supervision, Methodology, Conceptualization. **Ahmad Abbaszadeh:** Writing – review & editing, Validation, Supervision, Methodology, Conceptualization. **Wendy Novicoff:** Writing – review & editing, Validation, Supervision, Methodology, Investigation, Formal analysis, Data curation, Conceptualization.

## References

- [1] Wittig U, Moshhammer M, Vielgut I, Hauer G, Reinbacher P, Leithner A, et al. Higher use of fixed-bearing over mobile-bearing and posterior-stabilized over medial pivot designs in total knee arthroplasty (TKA): a systematic comparative analysis using worldwide arthroplasty registers from England and Wales, Australia, Norway, New Zealand, Germany and Switzerland. *Arch Orthop Trauma Surg* 2023;143:1021–9. <https://doi.org/10.1007/s00402-022-04410-8>.
- [2] Capella M, Dolfin M, Saccia F. Mobile bearing and fixed bearing total knee arthroplasty. *Ann Transl Med* 2016;4:127. <https://doi.org/10.21037/atm.2015.12.64>.
- [3] Abdel MP, Tibbo ME, Stuart MJ, Trousdale RT, Hanssen AD, Pagnano MW. A randomized controlled trial of fixed- versus mobile-bearing total knee arthroplasty: a follow-up at a mean of ten years. *Bone Joint J* 2018;100-B: 925–9. <https://doi.org/10.1302/0301-620X.100B7.BJJ-2017-1473.R1>.
- [4] Powell AJ, Crua E, Chong BC, Gordon R, McAuslan A, Pitto RP, et al. A randomized prospective study comparing mobile-bearing against fixed-bearing PFC Sigma cruciate-retaining total knee arthroplasties with ten-year

- minimum follow-up. *Bone Joint J* 2018;100-B:1336–44. <https://doi.org/10.1302/0301-620X.100B10.BJJ-2017-1450.R1>.
- [5] Goodfellow J, O'Connor J. The mechanics of the knee and prosthesis design. *J Bone Joint Surg Br* 1978;60-B:358–69. <https://doi.org/10.1302/0301-620X.60B3.581081>.
  - [6] Fisher J, McEwen H, Tipper J, Jennings L, Farrar R, Stone M, et al. Wear-simulation analysis of rotating-platform mobile-bearing knees. *Orthopedics* 2006;29:S36–41.
  - [7] Stukenborg-Colsman C, Ostermeier S, Hurschler C, Wirth CJ. Tibiofemoral contact stress after total knee arthroplasty: comparison of fixed and mobile-bearing inlay designs. *Acta Orthop Scand* 2002;73:638–46. <https://doi.org/10.1080/000164702321039598>.
  - [8] Poirier N, Graf P, Dubrana F. Mobile-bearing versus fixed-bearing total knee implants. Results of a series of 100 randomised cases after 9 years follow-up. *J Orthop Traumatol Surg Res* 2015;101:S187–92. <https://doi.org/10.1016/j.jotsr.2015.03.004>.
  - [9] Collier JP, Mayor MB, McNamara JL, Surprenant VA, Jensen RE. Analysis of the failure of 122 polyethylene inserts from uncemented tibial knee components. *Clin Orthop Relat Res* 1991;232–42.
  - [10] Bae JH, Kim JG, Lee SY, Lim HC, In Y, MUKA Study Group. Epidemiology of bearing dislocations after mobile-bearing Unicompartmental knee arthroplasty: multicenter analysis of 67 bearing dislocations. *J Arthroplasty* 2020;35:265–71. <https://doi.org/10.1016/j.arth.2019.08.004>.
  - [11] Hao D, Wang J. Fixed-bearing vs mobile-bearing prostheses for total knee arthroplasty after approximately 10 years of follow-up: a meta-analysis. *J Orthop Surg Res* 2021;16:437. <https://doi.org/10.1186/s13018-021-02560-w>.
  - [12] Hantouly AT, Ahmed AF, Alzobi O, Toubasi A, Salameh M, Elmhiregh A, et al. Mobile-bearing versus fixed-bearing total knee arthroplasty: a meta-analysis of randomized controlled trials. *Eur J Orthop Surg Traumatol* 2022;32:481–95. <https://doi.org/10.1007/s00590-021-02999-x>.
  - [13] Migliorini F, Maffulli N, Cuzzo F, Pilone M, Elsnar K, Eschweiler J. No difference between mobile and fixed bearing in primary total knee arthroplasty: a meta-analysis. *Knee Surg Sports Traumatol Arthrosc* 2022;30:3138–54. <https://doi.org/10.1007/s00167-022-07065-5>.
  - [14] Hofstede SN, Nouta KA, Jacobs W, van Hooff ML, Wymenga AB, Pijls BG, et al. Mobile bearing vs fixed bearing prostheses for posterior cruciate retaining total knee arthroplasty for postoperative functional status in patients with osteoarthritis and rheumatoid arthritis. *Cochrane Database Syst Rev* 2015;2015:CD003130. <https://doi.org/10.1002/14651858.CD003130.pub3>.
  - [15] Bakircioglu S, Aksoy T, Caglar O, Mazhar Tokgozoglul A, Atila B. Joint awareness after fixed and mobile-bearing total knee arthroplasty with minimum 12 years of follow-up: a propensity matched-pair analysis. *Knee* 2023;42:130–5. <https://doi.org/10.1016/j.knee.2023.03.004>.
  - [16] Sohn S, Koh JJ, Kim MS, Choi KY, Lim DS, In Y. Mobile-bearing has no benefit over fixed-bearing total knee arthroplasty in joint awareness and Crepitus: a randomized controlled trial. *J Arthroplasty* 2023;38:78–84. <https://doi.org/10.1016/j.arth.2022.07.021>.
  - [17] Radetzki F, Zeh A, Delank K-S, Wohlrab D. The high flex total knee arthroplasty-higher Incidence of aseptic loosening and No benefit in comparison to Conventional total knee arthroplasty: minimum 16-years follow-up results. *Indian J Orthop* 2021;55:76–80. <https://doi.org/10.1007/s43465-020-00276-3>.
  - [18] Kalaai S, Most J, van Dun B, Kaptein BL, Tilman PBJ, Boonen B, et al. Less wear in deep-dished mobile compared to fixed bearing total knee arthroplasty of the same design at 5-year follow-up: a randomised controlled model-based Roentgen stereophotogrammetric analysis trial. *Knee Surg Sports Traumatol Arthrosc* 2023;31:5137–44. <https://doi.org/10.1007/s00167-023-07584-9>.
  - [19] Page MJ, McKenzie JE, Bossuyt PM, Boutron I, Hoffmann TC, Mulrow CD, et al. The PRISMA 2020 statement: an updated guideline for reporting systematic reviews. *BMJ* 2021;372:n71. <https://doi.org/10.1136/bmj.n71>.
  - [20] Wylde V, Learmonth I, Potter A, Bettinson K, Lingard E. Patient-reported outcomes after fixed- versus mobile-bearing total knee replacement: a multicentre randomised controlled trial using the Kinemax total knee replacement. *J Bone Joint Surg Br* 2008;90:1172–9. <https://doi.org/10.1302/0301-620X.90B9.21031>.
  - [21] Tiwari V, Meshram P, Park CK, Bansal V, Kim TK. New mobile-bearing TKA with unique ball and socket post-cam mechanism offers similar function and stability with better prosthesis fit and gap balancing compared to an established fixed-bearing prosthesis. *Knee Surg Sports Traumatol Arthrosc* 2019;27:2145–54. <https://doi.org/10.1007/s00167-019-05430-5>.
  - [22] Scuderi G, Hedden D, Maltry J, Traina S, Sheinkop M, Hartzband M. Early clinical results of a high-flexion, posterior-stabilized, mobile-bearing total knee arthroplasty. *J Arthroplasty* 2012;27:421–9. <https://doi.org/10.1016/j.arth.2011.06.011>.
  - [23] Amaro JT, Arliani GG, Astur DC, Debieux P, Kaleka CC, Cohen M. No difference between fixed- and mobile-bearing total knee arthroplasty in activities of daily living and pain: a randomized clinical trial. *Knee Surg Sports Traumatol Arthrosc* 2017;25:1692–6. <https://doi.org/10.1007/s00167-016-4106-1>.
  - [24] Ferguson KB, Bailey O, Anthony I, James PJ, Stother IG, M JGB. A prospective randomised study comparing rotating platform and fixed bearing total knee arthroplasty in a cruciate substituting design—outcomes at two year follow-up. *Nee* 2014;21:151–5. <https://doi.org/10.1016/j.knee.2013.09.007>.
  - [25] Tienboon P, Jaruwangsanti N, Laohasinnurak P. A prospective study comparing mobile-bearing versus fixed-bearing type in total knee arthroplasty using the free-hand-cutting technique. *J Med Assoc Thai* 2012;95(Suppl 10):S77–86.
  - [26] Watanabe T, Tomita T, Fujii M, Hashimoto J, Sugamoto K, Yoshikawa H. Comparison between mobile-bearing and fixed-bearing knees in bilateral total knee replacements. *Int Orthop* 2005;29:179–81. <https://doi.org/10.1007/s00264-005-0646-6>.
  - [27] Choi W, Lee S, Seong S, Jung J, Lee M. Comparison between Standard and high-flexion posterior-stabilized rotating-platform mobile-bearing total knee arthroplasties A randomized controlled study. *J Bone Jt Surg Am Vol* 2010;92A:2634–42. <https://doi.org/10.2106/JBJS.L.01122>.
  - [28] Aglietti P, Baldini A, Buzzi R, Lup D, De Luca L. Comparison of mobile-bearing and fixed-bearing total knee arthroplasty: a prospective randomized study. *J Arthroplasty* 2005;20:145–53. <https://doi.org/10.1016/j.arth.2004.09.032>.
  - [29] Kim TK, Chang CB, Kang YG, Chung BJ, Cho HJ, Seong SC. Early clinical outcomes of floating platform mobile-bearing TKA: longitudinal comparison with fixed-bearing TKA. *Knee Surg Sports Traumatol Arthrosc* 2010;18:879–88. <https://doi.org/10.1007/s00167-009-0985-8>.
  - [30] Ball ST, Sanchez HB, Mahoney OM, Schmalzried TP. Fixed versus rotating platform total knee arthroplasty: a prospective, randomized, single-blind study. *J Arthroplasty* 2011;26:531–6. <https://doi.org/10.1016/j.arth.2010.06.003>.
  - [31] Harrington MA, Hopkinson WJ, Hsu P, Manion L. Fixed- vs mobile-bearing total knee arthroplasty: does it make a difference?—a prospective randomized study. *J Arthroplasty* 2009;24:24–7. <https://doi.org/10.1016/j.arth.2009.04.031>.
  - [32] Beard DJ, Pandit H, Price AJ, Butler-Manuel PA, Dodd CA, Murray DW, et al. Introduction of a new mobile-bearing total knee prosthesis: minimum three year follow-up of an RCT comparing it with a fixed-bearing device. *Knee* 2007;14:448–51. <https://doi.org/10.1016/j.knee.2007.08.004>.
  - [33] Luna JT, Sembrano JN, Gioe TJ. Mobile and fixed-bearing (all-polyethylene tibial component) total knee arthroplasty designs: surgical technique. *J Bone Joint Surg* 2010;92:240–9. <https://doi.org/10.2106/JBJS.J.00157>.
  - [34] Gioe TJ, Glynn J, Sembrano J, Suthers K, Santos ER, Singh J. Mobile and fixed-bearing (all-polyethylene tibial component) total knee arthroplasty designs. A prospective randomized trial. *J Bone Joint Surg Am* 2009;91:2104–12. <https://doi.org/10.2106/JBJS.H.01442>.
  - [35] Bailey O, Ferguson K, Crawford E, James P, May PA, Brown S, et al. No clinical difference between fixed- and mobile-bearing cruciate-retaining total knee arthroplasty: a prospective randomized study. *Knee Surg Sports Traumatol Arthrosc* 2015;23:1653–9. <https://doi.org/10.1007/s00167-014-2877-9>.
  - [36] Marques CJ, Daniel S, Sufi-Siavach A, Lampe F. No differences in clinical outcomes between fixed- and mobile-bearing computer-assisted total knee arthroplasties and no correlations between navigation data and clinical scores. *Knee Surg Sports Traumatol Arthrosc* 2015;23:1660–8. <https://doi.org/10.1007/s00167-014-3127-x>.
  - [37] Artz NJ, Hassaballa MA, Robinson JR, Newman JH, Porteous AJ, Murray JR. Patient reported kneeling ability in fixed and mobile bearing knee arthroplasty. *J Arthroplasty* 2015;30:2159–63. <https://doi.org/10.1016/j.arth.2015.06.063>.
  - [38] Rahman WA, Garbuz DS, Masri BA. Randomized controlled trial of radiographic and patient-assessed outcomes following fixed versus rotating platform total knee arthroplasty. *J Arthroplasty* 2010;25:1201–8. <https://doi.org/10.1016/j.arth.2009.10.002>.
  - [39] Hasegawa M, Sudo A, Uchida A. Staged bilateral mobile-bearing and fixed-bearing total knee arthroplasty in the same patients: a prospective comparison of a posterior-stabilized prosthesis. *Knee Surg Sports Traumatol Arthrosc* 2009;17:237–43. <https://doi.org/10.1007/s00167-008-0662-3>.
  - [40] Woolson ST, Epstein NJ, Huddleston JL. Long-term comparison of mobile-bearing vs fixed-bearing total knee arthroplasty. *J Arthroplasty* 2011;26:1219–23. <https://doi.org/10.1016/j.arth.2011.01.014>.
  - [41] Shemshaki H, Dehghani M, Eshaghi MA, Esfahani MF. Fixed versus mobile weight-bearing prosthesis in total knee arthroplasty. *Knee Surg Sports Traumatol Arthrosc* 2012;20:2519–27. <https://doi.org/10.1007/s00167-012-1946-1>.
  - [42] Sappey-Mariniere E, Swan J, Maucourt-Boulch D, Batailler C, Malatray M, Neyret P, et al. No significant clinical and radiological differences between fixed versus mobile bearing total knee replacement using the same semi-constrained implant type: a randomized controlled trial with mean 10 years follow-up. *Knee Surg Sports Traumatol Arthrosc* 2022;30:603–11. <https://doi.org/10.1007/s00167-020-06346-1>.
  - [43] Wohlrab D, Hube R, Zeh A, Hein W. Clinical and radiological results of high flex total knee arthroplasty: a 5 year follow-up. *Arch Orthop Trauma Surg* 2009;129:21–4. <https://doi.org/10.1007/s00402-008-0665-z>.
  - [44] Novaretti JV, Amaro JT, Astur DC, Cavalcante E, Rodrigues AG, Debieux P, et al. Higher axial tibiofemoral rotation and functional outcomes with mobile-bearing total knee arthroplasty compared to fixedbearing at one- but not at two-year follow-up: a randomized clinical trial. *J Knee Surg* 2020;5:314. <https://doi.org/10.1136/jisakos-2020-final-selected-abstracts.1>.
  - [45] Lizaar-Utrilla A, Sanz-Reig J, Trigueros-Rentero MA. Greater satisfaction in older patients with a mobile-bearing compared with fixed-bearing total knee arthroplasty. *J Arthroplasty* 2012;27:207–12. <https://doi.org/10.1016/j.arth.2011.05.021>.
  - [46] Tieskus CO, Daniilidis K, Skwara A, Dierkes T, Rosenbaum D, Fuchs-Winkelmann S. Gait analysis and electromyography in fixed- and mobile-



- bearing total knee replacement: a prospective, comparative study. *Knee Surg Sports Traumatol Arthrosc* 2011;19:2052–9. <https://doi.org/10.1007/s00167-011-1540-y>.
- [47] Park CH, Kang SG, Bae DK, Song SJ. Mid-term clinical and radiological results do not differ between fixed- and mobile-bearing total knee arthroplasty using titanium-nitride-coated posterior-stabilized prostheses: a prospective randomized controlled trial. *Knee Surg Sports Traumatol Arthrosc* 2019;27:1165–73. <https://doi.org/10.1007/s00167-018-5095-z>.
  - [48] Bhan S, Malhotra R, Kiran EK, Shukla S, Bijjawara M. A comparison of fixed-bearing and mobile-bearing total knee arthroplasty at a minimum follow-up of 4.5 years. *J Bone Joint Surg Am* 2005;87:2290–6. <https://doi.org/10.2106/JBJS.D.02221>.
  - [49] Jolles BM, Grzesiak A, Eudier A, Dejnabadi H, Voracek C, Pichonnaz C, et al. A randomised controlled clinical trial and gait analysis of fixed- and mobile-bearing total knee replacements with a five-year follow-up. *J Bone Joint Surg Br* 2012;94:648–55. <https://doi.org/10.1302/0301-620X.94B5.27598>.
  - [50] Feczko PZ, Jutten LM, van Steyn MJ, Deckers P, Emans PJ, Arts JJ. Comparison of fixed and mobile-bearing total knee arthroplasty in terms of patellofemoral pain and function: a prospective, randomised, controlled trial. *BMC Musculoskelet Disord* 2017;18:279. <https://doi.org/10.1186/s12891-017-1635-9>.
  - [51] Kim YH, Kook HK, Kim JS. Comparison of fixed-bearing and mobile-bearing total knee arthroplasties. *Clin Orthop Relat Res* 2001;101–15. <https://doi.org/10.1097/00003086-200111000-00013>.
  - [52] Rohella D, Behera AK, Sahu H, Madhual A, Mall B, Panda KK, et al. Comparison of fixed-bearing prosthesis vs. Mobile-bearing prosthesis in total knee arthroplasty among Indian patients. *Int J Pharmaceut Clin Res* 2022;14:7–12.
  - [53] Kim YH, Park JW, Kim JS. Comparison of high-flexion fixed-bearing and high-flexion mobile-bearing total knee arthroplasties-A prospective randomized study. *J Arthroplasty* 2018;33:130–5. <https://doi.org/10.1016/j.arth.2017.07.025>.
  - [54] Pijls BG, Valstar ER, Kaptein BL, Nelissen RGHH. Differences in long-term fixation between mobile-bearing and fixed-bearing knee prostheses at ten to 12 years' follow-up: a single-blinded randomised controlled radiostereometric trial. *J Bone Joint Surg Br* 2012;94 B:1366–71. <https://doi.org/10.1302/0301-620X.94B10.28858>.
  - [55] Breeman S, Campbell MK, Dakin H, Fiddian N, Fitzpatrick R, Grant A, et al. Five-year results of a randomised controlled trial comparing mobile and fixed bearings in total knee replacement. *Bone Joint J* 2013;95-B:486–92. <https://doi.org/10.1302/0301-620X.95B4.29454>.
  - [56] Vasdev A, Kumar S, Chadha G, Mandal SP. Fixed- versus mobile-bearing total knee arthroplasty in Indian patients. *J Orthop Surg* 2009;17:179–82. <https://doi.org/10.1177/230949900901700211>.
  - [57] Lädermann A, Lübbecke A, Stern R, Riand N, Fritschy D. Fixed-bearing versus mobile-bearing total knee arthroplasty: a prospective randomised, clinical and radiological study with mid-term results at 7 years. *Knee* 2008;15:206–10. <https://doi.org/10.1016/j.knee.2008.01.010>.
  - [58] Kim Y, Park J, Kim J, Kulkarni S. Long-term clinical outcomes and survivorship of Press-fit Condylar Sigma fixed-bearing and mobile-bearing total knee prostheses in the same patients. *J Bone Jt Surg Am Vol* 2014;96A:e168. <https://doi.org/10.2106/JBJS.M.01130>.
  - [59] Kim YH, Kim JS, Choe JW, Kim HJ. Long-term comparison of fixed-bearing and mobile-bearing total knee replacements in patients younger than fifty-one years of age with osteoarthritis. *J Bone Joint Surg Am* 2012;94:866–73. <https://doi.org/10.2106/JBJS.K.00884>.
  - [60] Killen CJ, Murphy MP, Hopkinson WJ, Harrington MA, Adams WH, Rees HW. Minimum twelve-year follow-up of fixed- vs mobile-bearing total knee arthroplasty: double blinded randomized trial. *J Clin Orthop Trauma* 2020;11:154–9. <https://doi.org/10.1016/j.jcot.2019.03.019>.
  - [61] Aggarwal AK, Agrawal A. Mobile vs fixed-bearing total knee arthroplasty performed by a single surgeon: a 4- to 6.5-year randomized, prospective, controlled, double-blinded study. *J Arthroplasty* 2013;28:1712–6. <https://doi.org/10.1016/j.arth.2013.01.003>.
  - [62] Baktir A, Karaaslan F, Yurdakul E, Karaoglu S. Mobile- versus fixed-bearing total knee arthroplasty: a prospective randomized controlled trial featuring 6–10-year follow-up. *Acta Orthop Traumatol Turc* 2016;50:1–9. <https://doi.org/10.3944/AOTT.2016.15.0120>.
  - [63] Matsuda S, Mizu-uchi H, Fukagawa S, Miura H, Okazaki K, Matsuda H, et al. Mobile-bearing prosthesis did not improve mid-term clinical results of total knee arthroplasty. *Knee Surg Sports Traumatol Arthrosc* 2010;18:1311–6. <https://doi.org/10.1007/s00167-010-1143-z>.
  - [64] Breugem SJ, van Ooij B, Haverkamp D, Siersevelt IN, van Dijk CN. No difference in anterior knee pain between a fixed and a mobile posterior stabilized total knee arthroplasty after 7.9 years. *Knee Surg Sports Traumatol Arthrosc* 2014;22:509–16. <https://doi.org/10.1007/s00167-012-2281-2>.
  - [65] Kalisvaart MM, Pagnano MW, Trousdale RT, Stuart MJ, Hanssen AD. Randomized clinical trial of rotating-platform and fixed-bearing total knee arthroplasty: no clinically detectable differences at five years. *J Bone Joint Surg Am* 2012;94:481–9. <https://doi.org/10.2106/JBJS.K.00315>.
  - [66] Mahoney OM, Kinsey TL, D'Errico TJ, Shen J. The John Insall Award: no functional advantage of a mobile bearing posterior stabilized TKA. *Clin Orthop Relat Res* 2012;470:33–44. <https://doi.org/10.1007/s11999-011-2114-5>.
  - [67] Nieuwenhuijse MJ, van der Voort P, Kaptein BL, van der Linden-van der Zwaag HM, Valstar ER, Nelissen RG. Fixation of high-flexion total knee prostheses: five-year follow-up results of a four-arm randomized controlled clinical and roentgen stereophotogrammetric analysis study. *J Bone Joint Surg Am* 2013;95:e1411–11. <https://doi.org/10.2106/JBJS.L.01523>.
  - [68] Van Hamersveld KT, Marang-Van De Mheen PJ, Van Der Heide HJL, Van Der Linden-Van Der Zwaag HMJ, Valstar ER, Nelissen RGHH. Migration and clinical outcome of mobile-bearing versus fixed-bearing single-radius total knee arthroplasty: a randomized controlled trial. *Acta Orthop* 2018;89:190–6. <https://doi.org/10.1080/17453674.2018.1429108>.
  - [69] Schotanus MGM, Pilot P, Kaptein BL, Draijer WF, Tilman PBJ, Vos R, et al. No difference in terms of radiostereometric analysis between fixed- and mobile-bearing total knee arthroplasty: a randomized, single-blind, controlled trial. *Knee Surg Sports Traumatol Arthrosc* 2017;25:2978–85. <https://doi.org/10.1007/s00167-016-4138-6>.
  - [70] Hansson U, Toksvig-Larsen S, Jörn LP, Ryd L. Mobile vs. fixed meniscal bearing in total knee replacement: a randomised radiostereometric study. *Knee* 2005;12:414–8. <https://doi.org/10.1016/j.knee.2004.12.002>.
  - [71] Kim YH, Park JW, Jang YS. Long-term (up to 27 Years) prospective, randomized study of mobile-bearing and fixed-bearing total knee arthroplasties in patients <60 Years of age with osteoarthritis. *J Arthroplasty* 2021;36:1330–5. <https://doi.org/10.1016/j.arth.2020.10.050>.
  - [72] Kim YH, Park JW, Kim JS. The long-term results of Simultaneous high-flexion mobile-bearing and fixed-bearing total knee arthroplasties performed in the same patients. *J Arthroplasty* 2019;34:501–7. <https://doi.org/10.1016/j.arth.2018.11.007>.
  - [73] Tjørnild M, Søballe K, Møller Hansen P, Holm C, Stilling M. Mobile-vs. Fixed-bearing total knee replacement: a randomized radiostereometric and bone mineral density study. *Acta Orthop* 2015;86:208–14. <https://doi.org/10.3109/17453674.2014.968476>.
  - [74] Chaudhry A, Goyal VK. Fixed-bearing versus high-flexion RP total knee arthroplasty (TKA): midterm results of a randomized controlled trial. *J Orthop Traumatol* 2018;19:2. <https://doi.org/10.1186/s10195-018-0493-z>.
  - [75] Radetzkij F, Wienke A, Mendel T, Gutteck N, Delank KS, Wohlrab D. High flex total knee arthroplasty—a prospective, randomized study with results after 10 years. *Acta Orthop Belg* 2013;79:536–40.
  - [76] Kim YH, Kim DY, Kim JS. Simultaneous mobile- and fixed-bearing total knee replacement in the same patients. A prospective comparison of mid-term outcomes using a similar design of prosthesis. *J Bone Joint Surg Br* 2007;89:904–10. <https://doi.org/10.1302/0301-620X.89B7.18635>.
  - [77] Fransen BL, Hoozemans MJ, Keijser LC, van Lent ME, Verheyen CC, Burger BJ. Does insert type affect clinical and functional outcome in total knee arthroplasty? A randomised controlled clinical trial with 5-year follow-up. *J Arthroplasty* 2015;30:1931–7. <https://doi.org/10.1016/j.arth.2015.05.018>.