MEDWATCH FDA Form 3500A, Mandatory Reporting

DRAFT for human review

Vigil decision: REPORTABLE (DEATH), confidence 95% | Priority: URGENT

A. PATIENT INFORMATION	
1. Patient identifier	UNKNOWN
2. Age	UNKNOWN
3. Sex	UNKNOWN
4. Weight	UNKNOWN

B. ADVERSE EVENT OR PRODUCT PROBLEM	
1. Check all that apply	[X] Adverse event [X] Product problem
2. Outcome attributed to adverse event	[X] Death [ ] Life-threatening [ ] Hospitalization (initial or prolonged) [ ] Disability or permanent damage [ ] Congenital anomaly/birth defect [ ] Required intervention to prevent permanent impairment/damage [ ] Other serious (important medical event)
3. Date of event	2020-12-28
4. Date of this report	2026-10-03
5. Describe event or problem	The manufacturer received information alleging that a Continuous Positive Airway Pressure (CPAP) device's sound abatement foam became degraded and caused a patient to develop lung cancer. The patient expired. This issue was reported to the FDA per 21 CFR 806. The device will be corrected per Res 88058.

D. SUSPECT MEDICAL DEVICE	
1. Brand name	DREAMSTATION AUTO CPAP
2a. Common device name	VENTILATOR, NON-CONTINUOUS (RESPIRATOR)
2b. Procode	BZD
3. Manufacturer name	PHILIPS RESPIRONICS, INC.
4. Model # / Lot # / Serial #	Model: UNKNOWN Lot: UNKNOWN Serial: UNKNOWN

G. ALL MANUFACTURERS	
4. Date received by manufacturer	2021-10-15
7. Type of report	[ ] 5-day [X] 30-day

H. DEVICE MANUFACTURERS ONLY	
1. Type of reportable event	[X] Death [ ] Serious injury [ ] Malfunction
3. Device evaluated by manufacturer?	[ ] Yes [ ] No [X] Unknown
10. Additional manufacturer narrative	The manufacturer received information alleging that a Continuous Positive Airway Pressure (CPAP) device's sound abatement foam became degraded and caused a patient to develop lung cancer. The patient expired. This issue was reported to the FDA per 21 CFR 806. The device will be corrected per Res 88058.

H. DEVICE MANUFACTURERS ONLY	
10. Follow-up questions	1. Patient age and sex 2. Patient weight 3. Specific details on the nature of the foam degradation 4. Causal link between foam degradation and lung cancer 5. Device serial number 6. Device lot/batch number 7. Information on whether the device was returned for evaluation

Fields marked UNKNOWN were not stated in the complaint and must be obtained before submission. This draft is not an FDA submission; a qualified reviewer must verify every field.