
true force in the specimen (simulation only)
force read by the output gauge (measured in the test; one colour per rate)
UT19, 100 s^{-1}, centred (gaps)
max gauge error after 5 %: 3 %
UT19, 500 s^{-1}, centred (gaps)
max gauge error after 5 %: 15 %
UT19, 1000 s^{-1}, centred (gaps)
max gauge error after 5 %: 39 %
UT19, 100 s^{-1}, in contact
max gauge error after 5 %: 1 %
UT19, 500 s^{-1}, in contact
max gauge error after 5 %: 4 %
UT19, 1000 s^{-1}, in contact
max gauge error after 5 %: 5 %
force (kN)
engineering strain (%)
0
5
10
15
20
25
30