



# SHEET NOTES

1. VBSE PROVIDED BY FLOW SCIENCES. REFER TO MANUAL FOR OPERATING INSTRUCTIONS. NO ADDITIONAL EQUIPMENT NEEDED FOR INDEPENDENT RECIRCULATION.
2. SPENCER VALVE FITTING.

PROJECT LOCATION

6701 EVANSTAD DR N  
MAPLE GROVE, MN 55369



CLIENT

**UPSHER-SMITH**

Partners in Health Since 1919

CONSULTING ENGINEER

**Horwitz**

7400 49th AVENUE NORTH  
NEW HOPE, MN 55428  
(763) 533-1900  
www.horwitzinc.com

PROJECT TITLE

PROJECT KEystone - PHASE 1

REV	BY	DATE	DESCRIPTION
0	RS	05/25/22	AS RECORDED

LICENSEE

I HEREBY CERTIFY THAT THIS PLAN, SPECIFICATION OR REPORT WAS PREPARED BY ME OR UNDER MY DIRECT SUPERVISION AND THAT I AM A DULY LICENSED ENGINEER UNDER THE LAWS OF THE STATE OF MINNESOTA.

Date 05/25/22 LIC NO

CONFIDENTIAL - THIS DRAWING IS THE PROPERTY OF UPSHER-SMITH. THE INFORMATION CONTAINED HEREIN MAY NOT BE USED OR COPIED IN ANY MANNER WITHOUT THE WRITTEN PERMISSION OF THE OWNER.

PROJ. NO.	2022.1720	DESIGNED BY:	Designer
SCALE:	12" = 1'-0"	CHECKED BY:	Checker
SHEET SIZE:	30x42	APPROVED BY:	Approver

DRAWING TITLE

WEIGHING ROOM 1625

DRAWING NUMBER

PID-1020

WD06-SC-001  
GRAM SCALE  
CAPACITY: 100 - 15KG  
VENDOR: METTLER TOLEDO  
MODEL: ICS466d-AB15

WD06-VBSE-001  
ASSEMBLY/SKID/BOUNDARY  
WD06-VBSE-001  
VENDOR: FLOW SCIENCES

WD06-MIL-001  
COMIL  
VENDOR: QUADRO

WD06-VS-001  
18" - 24" VIBRATORY SIEVE  
VENDOR: TBD

WD06-SC-002  
DECKHAND SCALE  
RANGE: 2 KG - 300 KG  
VENDOR: METTLER TOLEDO  
MODEL: 2888