

Calibration System Audit - 2026

Cascade Precision Components - measurement equipment audit

Date: 2026-05-16 Scope: all six plants, 412 gauges

Results: 407 of 412 instruments within calibration (98.8%).

Five overdue: three at Bend (plant Idle - tagged out of service), one torque wrench at Salem (recalled same day), one micrometer at Spokane (in transit to lab).

Gauge R&R studies current for all CMMs and final-inspection instruments. No product impact identified.

Overall: SATISFACTORY. Next audit 2027-05.

All Cascade Precision Components data is fictional and generated for training.